


**Tektronix**<sup>®</sup>  
COMMITTED TO EXCELLENCE



INSTRUCTION MANUAL

Tektronix, Inc.  
P.O. Box 500  
Beaverton, Oregon 97077

Serial Number \_\_\_\_\_

070-2726-01

First Printing JULY 1979



## WARRANTY

Tektronix warrants to the original purchaser that this product is free from defects in materials and workmanship under normal use for a period of one (1) year from the date of shipment. Tektronix will, at its option, repair or replace the product if Tektronix determines it is defective within the warranty period, and it is returned, freight prepaid, to a Tektronix Service Center.

There is no implied warranty of fitness for a particular purpose. Tektronix is not liable for consequential damages.

# TABLE OF CONTENTS

	Page
LIST OF ILLUSTRATIONS .....	iv
LIST OF TABLES .....	iv
OPERATOR'S SAFETY SUMMARY .....	v
SERVICE SAFETY SUMMARY .....	vii
<b>SECTION 1</b>	
<b>GENERAL INFORMATION AND SPECIFICATIONS</b>	
GENERAL INFORMATION .....	1-1
INTRODUCTION .....	1-1
STANDARDS, DOCUMENTS, AND REFERENCES USED .....	1-1
CHANGE AND HISTORY INFORMATION .....	1-1
PRODUCT DESCRIPTION .....	1-1
SPECIFICATION .....	1-2
ELECTRICAL CHARACTERISTICS .....	1-2
Frequency Related .....	1-2
Amplitude Related .....	1-5
Input Signal Characteristics .....	1-7
Output Signal Characteristics .....	1-8
General Characteristics .....	1-8
Power Requirements .....	1-8
ENVIRONMENTAL CHARACTERISTICS .....	1-9
PHYSICAL CHARACTERISTICS .....	1-10
ACCESSORIES .....	1-10
OPTIONS .....	1-12
Option 1 .....	1-12
Option 2 .....	1-14
Option 3 .....	1-14
Option 8 .....	1-15
Option 20 .....	1-16
Option 21 .....	1-16
Option 22 .....	1-16
Options for Power Cord Configuration .....	1-17
<b>SECTION 2</b>	
<b>INSTALLATION INSTRUCTIONS</b>	
INTRODUCTION .....	2-1
UNPACKING AND INITIAL INSPECTION .....	2-1
PREPARATION FOR USE .....	2-1
POWER SOURCE AND POWER REQUIREMENTS .....	2-2
REPACKAGING FOR SHIPMENT .....	2-2

# TABLE OF CONTENTS (cont)

SECTION 3	OPERATION	Page
	CONTROLS, INDICATORS, AND CONNECTORS .....	3-1
	TURN-ON AND PREPARATION FOR USE .....	3-7
	1. Initial Turn On .....	3-7
	2. Calibrate Center Frequency Readout .....	3-8
	3. Calibrate Reference Level and Dynamic Range .....	3-8
	4. Check Span Accuracy and Linearity .....	3-8
	FUNCTIONAL OR OPERATIONAL CHECK .....	3-9
	1. Check Operation of Front Panel Pushbuttons and Controls .....	3-9
	2. Check Frequency Readout Accuracy .....	3-11
	3. Check Frequency Span/Div Range and Accuracy .....	3-12
	4. Check Resolution Bandwidth and Shape Factor .....	3-13
	5. Check Reference Level Gain and RF Attenuator Steps .....	3-13
	6. Check Sensitivity .....	3-14
	7. Check Frequency Drift or Stability .....	3-16
	8. Check Residual FM .....	3-17
	9. Digital Storage (Option 2) .....	3-17
	GENERAL OPERATING INFORMATION .....	3-18
	Crt Light Filters .....	3-18
	Intensity Level, Focus, and Beam Alignment .....	3-18
	Signal Application .....	3-18
	RF INPUT Connector .....	3-18
	Amplitude Conversion .....	3-19
	Connecting to 75 $\Omega$ Source .....	3-19
	Resolution Bandwidth, Frequency Span, and Sweep Time .....	3-19
	Using the Peaking Control .....	3-20
	Using the Signal Identifier .....	3-21
	Using the Video Filters .....	3-21
	Time Domain Operation .....	3-21
	Triggering the Display .....	3-21
	Sweeping the Display .....	3-22
	Manual Scan of the Spectrum .....	3-22
	Reference Level, RF Attenuation, and Vertical Display .....	3-22
	Delta A Mode .....	3-22
	Min Noise/Min Distortion .....	3-23
	Digital Storage (Option 2) .....	3-23
	Waveguide Mixers and External Diplexer .....	3-24
	Handling .....	3-24
	Installation .....	3-24
	Operation .....	3-25
	Analyzing Signals .....	3-26

# TABLE OF CONTENTS (cont)

SECTION 3	OPERATION (cont)	Page
	OPERATIONAL PRECAUTIONS .....	3-26
	1. Measurements Outside the Specified Frequency and Tuning Range Versus Span of the Display .....	3-26
	2. Signal FM .....	3-26
	3. Correct Trigger Mode .....	3-26
	4. Level of Pulsed Signals .....	3-26
	5. Level of Continuous Wave Signals .....	3-27
	6. Excessive Input Signal Level .....	3-27
	7. No Crt Trace .....	3-27
	8. Digital Storage Effects on Signal Analyses .....	3-27
	9. Stored Display Averaged in Wide Spans .....	3-27
	10. Cold Storage or Power Interrupt Initialization .....	3-27
	SERVICE MANUALS .....	3-28
	PRODUCT SERVICE .....	3-28
SECTION 4	OPTION INFORMATION .....	4-1
APPENDIX A	GLOSSARY .....	A-1

## CHANGE INFORMATION

# LIST OF ILLUSTRATIONS

Figure		Page
1-1	Dimensions .....	1-10
1-2	International power and plug configuration for the 492 .....	1-17
3-1	Front panel selectors, controls, and connectors .....	3-1
3-2	Rear panel connectors .....	3-6
3-3	Crt readout for power-up state .....	3-7
3-4	Typical display of calibrator markers in MAX SPAN position .....	3-7
3-5	Displays that illustrate how bandwidth and shape factor are determined .....	3-13
3-6	Typical display using digital storage with MAX HOLD activated, to measure drift .	3-16
3-7	Displays to illustrate how residual FM is measured .....	3-17
3-8	Using digital storage to measure the differential between two events .....	3-18
3-9	Volts-dBm-Watts conversion chart for 50 $\Omega$ impedance .....	3-19
3-10	Circuit of a 75 $\Omega$ to 50 $\Omega$ matching pad (ac coupled) .....	3-20
3-11	Graph to illustrate the relationship between dBm, dBmV, and dB $\mu$ V (matching attenuator included where necessary) .....	3-20
3-12	Integrating the display with the Video Filter .....	3-21
3-13	External (Waveguide) mixer installation .....	3-24
3-14	Typical display generated by a signal into the waveguide mixer .....	3-25
3-15	Identifier mode displays .....	3-26

# LIST OF TABLES

Table		Page
1-1	Frequency Response and Flatness .....	1-4
1-2	Sensitivity versus Bandwidth .....	1-6
1-3	Accessories .....	1-10
1-4	Option 1 Frequency Characteristics .....	1-12
1-5	Option 1 Sensitivity .....	1-13
2-1	Shipping Carton Test Strength .....	2-2
3-1	Span/Div Ranges versus Band and Option .....	3-3
3-2	Narrow and Wide Spans versus Frequency Band .....	3-12
3-3	Sensitivity without Preselector .....	3-15
3-4	Sensitivity Option 1 .....	3-16

# OPERATORS SAFETY SUMMARY

The general safety information in this part of the summary is for both operating and servicing personnel. Specific warnings and cautions will be found throughout the manual where they apply, but may not appear in this summary.

## TERMS

### In This Manual

CAUTION statements identify conditions or practices that could result in damage to the equipment or other property.

WARNING statements identify conditions or practices that could result in personal injury or loss of life.

### As Marked on Equipment

CAUTION indicates a personal injury hazard not immediately accessible as one reads the marking, or a hazard to property including the equipment itself.

DANGER indicates a personal injury hazard immediately accessible as one reads the marking.

## SYMBOLS

### In This Manual



This symbol indicates where applicable cautionary or other information is to be found.

### As Marked on Equipment



DANGER — High voltage.



Protective ground (earth) terminal.



ATTENTION — refer to manual.

### Power Source

This product is intended to operate from a power source that will not apply more than 250 volts rms between the supply conductors or between either supply conductor and ground. A protective ground connection by way of the grounding conductor in the power cord is essential for safe operation.

### **Grounding the Product**

This product is grounded through the grounding conductor of the power cord. To avoid electrical shock, plug the power cord into a properly wired receptacle before connecting to the product input or output terminals. A protective ground connection by way of the grounding conductor in the power cord is essential for safe operation.

### **Danger From Loss of Ground**

Upon loss of the protective-ground connection, all accessible conductive parts (including knobs and controls that may appear to be insulating) can render an electric shock.

### **Use the Proper Power Cord**

Use only the power cord and connector specified for your product.

Use only a power cord that is in good condition.

For detailed information on power cords and connectors, see Fig. 1-2.

Refer cord and connector changes to qualified service personnel.

### **Use the Proper Fuse**

To avoid fire hazard, use only the fuse of correct type, voltage rating and current rating as specified in the parts list for your product.

Refer fuse replacement to qualified service personnel.

### **Do Not Operate in Explosive Atmospheres**

To avoid explosion, do not operate this product in an explosive atmosphere unless it has been specifically certified for such operation.

### **Do Not Remove Covers or Panels**

To avoid personal injury, do not remove the product covers or panels. Do not operate the product without the covers and panels properly installed.



# SERVICE SAFETY SUMMARY

## FOR QUALIFIED SERVICE PERSONNEL ONLY

*Refer also to the preceding Operators Safety Summary.*

### **Do Not Service Alone**

Do not perform internal service or adjustment of this product unless another person capable of rendering first aid and resuscitation is present.


### **Use Care When Servicing With Power On**

Dangerous voltages exist at several points in this product. To avoid personal injury, do not touch exposed connections and components while power is on.

Disconnect power before removing protective panels, soldering, or replacing components.

### **Power Source**

This product is intended to operate from a power source that will not apply more than 250 volts rms between the supply conductors or between either supply conductor and ground. A protective ground connection by way of the grounding conductor in the power cord is essential for safe operation.



2726-12

492 Spectrum Analyzer.

# GENERAL INFORMATION AND SPECIFICATION

## GENERAL INFORMATION

### Introduction

This manual contains information pertinent to the installation and operation of the 492 Spectrum Analyzer. Contents and organization of the manual are described in the Table of Contents preceding this section. These instructions assume the user is knowledgeable in frequency domain analysis. The intent is to provide information necessary to effectively operate the 492. Service information is in a separate service manual.

### Standards, Documents, and References Used

Terminology used in the manual is in accordance with industry practice. Abbreviations are in accordance with ANSI Y1.1-1972, with exceptions and additions explained in parentheses after the abbreviation. Graphic symbology is based on ANSI Y32.2-1975. Logic symbology is based on ANSI Y32.14-1973 and the manufacturer's data books or sheets. A copy of ANSI standards may be obtained from the Institute of Electrical and Electronic Engineers, 345 47th Street, New York, NY 10017.

### Change and History Information

Change information that involves manual corrections and/or additional data is located at the back of the manual in the CHANGE INFORMATION section.

History information with the updated data is integrated into the text or diagrams when a page or diagram is updated.

### Product Description

The 492 Spectrum Analyzer is a high performance, compact, portable spectrum analyzer that displays absolute amplitude and frequency information of signals within the frequency spectrum of 50 kHz to 21 GHz with the internal coaxial mixer, and up to 60 GHz with optional external TEKTRONIX High Performance Waveguide Mixers. The 8.4 X 10.2 cm crt face reads out all major display parameters.

Some features of the 492 are: Simplified operation through the use of an internal microcomputer. Display dynamic range of 80 dB with calibrated reference level readout from  $-123$  dBm to  $+30$  dBm, in 10 dB and 1 dB steps. When using the Delta A mode, to measure the amplitude difference between two signals, the steps are 0.25 dB. Resolution bandwidths from 1 kHz to 1 MHz in decade steps, with a shape factor of 7.5:1 or better (options provide additional bandwidth selections of 100 Hz). Intermodulation products are: 70 dB or more down, harmonic distortion is  $-70$  dBc or better. Sensitivity of  $-115$  dBm to 7.1 GHz, at 1 kHz resolution bandwidth. Frequency response of  $\pm 1.5$  dB to 7.1 GHz and  $\pm 2.5$  dB to 18 GHz. Digital storage (with Option 2) features peak detection and digital signal averaging.

# SPECIFICATION

The following list of instrument characteristics and features apply to the basic 492 Spectrum Analyzer after a 30 minute warmup, except as noted. Changes to the basic specifications due to the addition of options follow the basic listings.

Performance requirements are provided in the Calibration section of the Service Instructions. The Performance Check procedures require sophisticated equipment as well as technical expertise to perform.

The Performance Requirement column describes the limits of the characteristic, and the Supplemental column describes features and typical values or information that may be useful to the user. Procedures to verify perfor-

mance requirements are provided in the Calibration section of the Service Instructions. The Performance Check procedures require sophisticated equipment as well as technical expertise to perform.

## ELECTRICAL CHARACTERISTICS

Characteristic	Performance Requirement	Supplemental Information
FREQUENCY RELATED		
Center Frequency		
Range (Internal Mixer)	100 kHz to 18 GHz.	Usable to 50 kHz and 21 GHz with reduced performance.
Accuracy (after 2 hour warmup)	$\pm(5 \text{ MHz} + 20\% \text{ of span/div})$ or $\pm(0.2\% \text{ of the center frequency} + 20\% \text{ of the span/div})$ which ever is greater.	
Readout Resolution (after 2 hour warmup)		Within 1 MHz.
Residual FM (short term)	$\leq 1 \text{ kHz peak-to-peak}$ $n$ for a period of 20 ms. $n$ is the 1st LO harmonic number used in the 1st mixer conversion, and related to the selected frequency range (band).	No video filter.
Frequency Drift	$\leq 200 \text{ kHz/hour}$ after 2 hour warmup.	
"Static" Resolution Bandwidth (6 dB down)	1 kHz to 1 MHz in decade steps, plus an automatic (AUTO position). Resolution bandwidth (6 dB down) is within 20% of the bandwidth selected.	In AUTO position the bandwidth is computed by an internal computer, based on the span/div, video filters, time/div, and vertical display selections. When both the TIME/DIV and RESOLUTION BANDWIDTH are in AUTO position, the resolution bandwidth is a function of the FREQUENCY SPAN/DIV selection.
Shape Factor (60 dB/6 dB)	7.5:1 or better.	

## ELECTRICAL CHARACTERISTICS (cont)

Characteristic	Performance Requirement	Supplemental Information		
Noise Sidebands	At least 75 dBc at 30 times the resolution offset (70 dBc for 1 kHz resolution bandwidth) for fundamental mixing.			
Video Filter Narrow		Reduces video bandwidth to approximately 1/300th of the selected resolution bandwidth.		
Wide		Reduces video bandwidth to approximately 1/30th of the selected resolution bandwidth.		
Frequency Span/Div Range		<b>Band</b>	<b>Narrow Span</b>	<b>Wide Span</b>
		1—3 (0—7.1 GHz)	10 kHz/Div	200 MHz/Div
		4—5 (5.4—21 GHz)	50 kHz/Div	500 MHz/Div
		6 (18—26 GHz)	50 kHz/Div	1 GHz/Div
		7—8 (26—60 GHz)	100 kHz/Div	2 GHz/Div
		9 (60—90 GHz)	200 kHz/Div	2 GHz/Div
		10 (90—140 GHz)	500 kHz/Div	5 GHz/Div
		11 (140—220 GHz)	500 kHz/Div	10 GHz/Div
		Two additional positions provide full band (MAX span) display or 0 Hz (time domain) display.		
Accuracy	Within 5% of the span/div selected over the center eight divisions of a ten division display.			
Frequency Response and Display Flatness	Refer to Table 1-1.	Frequency response is measured with 10 dB or more of RF attenuation and with the PEAKING optimized when applicable at each frequency setting. It includes the effects of input vswr, mixing mode (n), gain variation, pre-selector, and mixer. Display flatness is typically $\pm 1$ dB greater than the frequency response.		

## ELECTRICAL CHARACTERISTICS (cont)

Table 1-1

## FREQUENCY RESPONSE AND FLATNESS

Frequency Band	Frequency Range	LO Frequency & Harmonic (n) <sup>c</sup>	1st IF Frequency	Frequency Response	FREQ Response Referred to 100 MHz
WITH COAXIAL (direct) INPUT					
1	0—4.2 GHz <sup>b</sup>	(2072—6272) (1-)	2072 MHz	±2.5 dB (50 kHz—4.2 GHz) ±1.5 dB (100 kHz—4.2 GHz)	
2	1.7—5.5 GHz	(2529—6329) (1-)	829 MHz	±1.5 dB	±2.5 dB
3	3.0—7.1 GHz	(2171—6271) (1+)	829 MHz	±1.5 dB	±2.5 dB
4	5.4—18.0 GHz	(2072—6276) (3-)	829 MHz	±2.5 dB	±3.5 dB
5	15.0—21.0 GHz	(4309—6309) (3+)	2072 MHz	±3.5 dB	
WITH EXTERNAL WAVEGUIDE MIXERS					
6	18.0—26 GHz <sup>d</sup>	(2655—4071) (6+)	2072 MHz	±3.0 dB <sup>f</sup>	±6 dB
7	26—40.0 GHz <sup>d</sup>	(2443—3793) (10+)	2072 MHz	±3.0 dB <sup>f</sup>	±6 dB
8	40.0—60.0 GHz <sup>d</sup>	(3790—5790) (10+)	2072 MHz	±3.0 dB <sup>f</sup>	±6 dB
9	60.0—90.0 GHz <sup>e</sup>	(3861—5862) (15+)	2072 MHz	—	
10	90.0—140 GHz <sup>e</sup>	(3822—6000) (23+)	2072 MHz	—	
11	140—220 GHz <sup>e</sup>	(3728—5890) (37+)	2072 MHz	—	

<sup>a</sup> Frequency range of the coaxial input bands is 100 kHz to 18 GHz; usable with reduced performance, to 50 kHz and 21 GHz.

<sup>b</sup> Frequency range of non-option 1 instruments.

<sup>c</sup> "n" denotes the LO (local oscillator) harmonic number. The + or - sign indicate which side (above or below) the incoming signal is with respect to the effective LO harmonic.

<sup>d</sup> Can be tuned over wider frequency range with unspecified performance.

<sup>e</sup> Dependent on external mixer.

<sup>f</sup> With TEKTRONIX High Performance Waveguide Mixers.

## ELECTRICAL CHARACTERISTICS (cont)

Characteristics	Performance Requirements	Supplemental Information
AMPLITUDE RELATED		
Display Modes		10 dB/Div, 2 dB/Div, Linear, and Delta A.
Display Reference Level (full screen or top graticule line, in Log mode)		<p>-123 dBm to +40 dBm in calibrated 10 dB or 1 dB steps, for 10 dB/Div mode and 1 dB for the 2 dB/Div mode. (+40 dBm includes 10 dB of IF gain reduction. Maximum input level remains at +30 dBm.)</p> <p>In LIN mode, the range is 20 nV/Div to 500 mV/Div, MIN DISTORTION mode. With Reference Level in FINE step function, the LIN mode increments (steps) are the equivalent of 1 dB.</p>
Accuracy		Function of the RF attenuation and reference level settings, resolution switching, frequency response, frequency band, and display mode. See amplitude accuracies of these functions. The attenuator is changed for reference levels above -30 dBm (-20 dBm in minimum noise) and IF gain is added for reference levels below -30 dBm (-20 dBm in minimum noise) unless a MIN RF ATTEN setting greater than the nominal attenuation is specified.
Display Dynamic Range		80 dB at 10 dB/Div and 16 dB at 2 dB/Div.
Accuracy	±1.0 dB/10 dB to a maximum cumulative error of ±2.0 dB over the 80 dB window and ±0.4 dB/2 dB to a maximum cumulative error of ±1.0 dB over the 16 dB window. LIN mode is ±5% of full.	
RF Attenuator		
Range		0 to 60 dB in 10 dB steps.
Accuracy		
DC to 4 GHz	Within 0.3 dB/10 dB to a maximum of 0.7 dB over the 60 dB range.	
4 GHz to 18 GHz	Within 0.5 dB/10 dB to a maximum of 1.4 dB over the 60 dB range.	
IF Gain		
Range		83 dB of gain increase, 10 dB of gain decrease (MIN NOISE activated) in 10 dB and 1 dB steps.
Accuracy	Within 0.2 dB/dB and 0.5 dB/10 dB, to a maximum of ±2 dB over the 90 dB range.	

ELECTRICAL CHARACTERISTICS (cont)

Characteristic	Performance Requirement	Supplemental Information															
Differential Amplitude (Delta A mode)																	
Measurement																	
Accuracy	<table border="1"> <thead> <tr> <th>dB Difference</th> <th>Steps</th> <th>Error</th> </tr> </thead> <tbody> <tr> <td>0.25 dB</td> <td>1</td> <td>0.05 dB</td> </tr> <tr> <td>2 dB</td> <td>8</td> <td>0.4 dB</td> </tr> <tr> <td>10 dB</td> <td>40</td> <td>1.0 dB</td> </tr> <tr> <td>50 dB</td> <td>200</td> <td>2.0 dB</td> </tr> </tbody> </table>	dB Difference	Steps	Error	0.25 dB	1	0.05 dB	2 dB	8	0.4 dB	10 dB	40	1.0 dB	50 dB	200	2.0 dB	Within the reference level range of -123 dBm to +30 dBm.
dB Difference	Steps	Error															
0.25 dB	1	0.05 dB															
2 dB	8	0.4 dB															
10 dB	40	1.0 dB															
50 dB	200	2.0 dB															
Range	From 10 dB above to 40 dB below the reference level established when the Delta A mode was activated.	Do not use Delta A mode outside the -123 dBm to +30 dBm reference level range. Total range is at least 50 dB.															
Signal Amplitude Variation With Resolution Switching	Less than 0.5 dB.																
Sensitivity																	

The following tabulation shows the equivalent maximum input noise for each resolution bandwidth, with the internal mixer for frequency bands 1—5 (100 kHz—18 GHz), and TEKTRONIX High Performance Waveguide Mixers for bands 6—10 (18 GHz—140 GHz). The NARROW video filter is activated, for narrow resolutions (1 kHz or less); WIDE filter for wide resolution.

Table 1-2  
SENSITIVITY VERSUS BANDWIDTH

Frequency/Band	Equivalent Input Noise for Resolution Bandwidths			
	1 kHz	10 kHz	100 kHz	1 MHz
50 kHz—7.1 GHz (Bands 1—3)	-115 dBm	-105 dBm	-95 dBm	-85 dBm
5.4—18.0 GHz (Band 4)	-100 dBm	-90 dBm	-80 dBm	-70 dBm
15—21.0 GHz (Band 5)	-95 dBm	-85 dBm	-75 dBm	-65 dBm
18.0—26 GHz (Band 6) <sup>a</sup>	-100 dBm	-90 dBm	-80 dBm	-70 dBm
26—40.0 GHz (Band 7) <sup>a</sup>	-95 dBm	-85 dBm	-75 dBm	-65 dBm
40.0—60.0 GHz (Band 8) <sup>a</sup>	-95 dBm	-85 dBm	-75 dBm	-65 dBm
60.0—90.0 GHz (Band 9)	External Mixer Dependent			
90.0—140 GHz (Band 10)	External Mixer Dependent			
140—220 GHz (Band 11)	External Mixer Dependent			

<sup>a</sup> TEKTRONIX High Performance Waveguide Mixers.



## ELECTRICAL CHARACTERISTICS (cont)

Characteristics	Performance Requirements	Supplemental Information
Spurious Response		
Residual (no input signal, referenced to mixer input, and fundamental mixing for bands 1, 2, and 3)	-100 dBm or less.	
Third order intermodulation products (MIN DISTORTION mode)	At least -70 dBc below any two on screen signals within any frequency span.	
Harmonic Distortion (cw signal, MIN DISTORTION mode)	At least -60 dBc for full screen signal.	
LO Emissions (referenced to input mixer)	-10 dBm or less.	

## INPUT SIGNAL CHARACTERISTICS

RF INPUT		Type N female connector, specified to 18 GHz, usable to 21 GHz.
Input Impedance		50 ohm; vswr 1:45 maximum, with 10 dB or more RF attenuation, to 18 GHz and 3.5 to 21 GHz.
Input Level		
Optimum level for linear operation	-30 dBm, referenced to input mixer.	This is achieved by being in MIN DISTORTION and not exceeding full screen.
1 dB compression point	-10 dBm, no RF attenuation.	
Maximum input level		
RF Attenuation at 0 dB	+13 dBm (Input mixer limit).	
With 20 dB or more RF Attenuation	+30 dBm (1 w) continuous, 75 w peak, pulse width 1 $\mu$ s or less with a maximum duty factor of 0.001 (attenuator limit).	Do Not apply dc voltage to the RF INPUT.
External Mixer		Input for IF signal and the source of negative-going bias for external waveguide mixers. Bias range +0.5 to -1.5 V.
EXT IN HORIZ/TRIG		Dc coupled input for horizontal drive and ac coupled for trigger signal.
Input Voltage Range		
Sweep		0 to +10 V (dc + peak ac) for full screen deflection.
Trigger	0.5 V peak (minimum) to 50 V peak (maximum). Pulse width 0.1 $\mu$ s minimum. Frequency 15 Hz to 1 MHz.	
ACCESSORY (J104)		This connector may be used for future applications.

## ELECTRICAL CHARACTERISTICS (cont)

Characteristics	Performance Requirements	Supplemental Information
OUTPUT SIGNAL CHARACTERISTICS		
Calibrator (CAL OUT)	-20 dBm $\pm$ 0.3 dB at 100 MHz $\pm$ 0.01%.	100 MHz comb markers are provided for frequency and span calibration.
1st LO and 2nd LO		Provides access to the output of the respective local oscillators (1st LO +7.5 dBm minimum, 2nd LO -16 dBm minimum). These ports must be terminated in 50 $\Omega$ at all times.
Vertical	Provides 0.5 V $\pm$ 5% of signal per division of video above and below the centerline. Centerline represents 0 V.	Source impedance approximately 1 k $\Omega$ .
Horiz Out	Provides 1.0 V/Div either side of center. Full range -2.5 V to +2.5 V $\pm$ 10%.	Source impedance approximately 1k $\Omega$ .
Pen Lift		TTL compatible, nominal +5 volts to lift pen.
IF Out		Access to the 10 MHz IF. Output level is approximately -15 dBm for a full screen signal at -30 dBm level reference. Nominal impedance approximately 50 $\Omega$ .

## GENERAL CHARACTERISTICS

Sweep		Triggered, auto, manual, and external.
Sweep Time	20 $\mu$ s/Div to 5 s/Div in 1-2-5 sequence (10 s/Div in Auto).	
Accuracy	$\pm$ 5%.	
Triggering	$\geq$ 2.0 division of signal for internal, and 0.5 V peak minimum to 50 V peak maximum external.	Internal, external, free run, and single sweep. Internal is ac coupled (15 Hz to 1 MHz).
Crt Readout		Displays: Reference level, frequency, vertical display mode, frequency span/div, frequency range, resolution bandwidth, and RF attenuation.

## POWER REQUIREMENTS

Characteristic	Description
Input Voltage	90 to 132 Vac or 180 to 250 Vac, 48 to 440 Hz.
Power (Options 1, 2, 3)	At 115 V, 60 Hz; 210 watts maximum, 3.2 amperes.
Leakage Current	5 mA peak.

## NOTE

If power to this instrument is interrupted, it may be necessary to re-initialize the microcomputer; when power is restored, turn the POWER switch Off for 5 seconds then back On.

## NOTE

After storage at temperatures below the operating range, the microcomputer may not initialize on power-up. If so, allow the instrument to warm up for 15 minutes and re-initialize the microcomputer by turning the POWER Off for 5 seconds then back On.

## ENVIRONMENTAL CHARACTERISTICS

Meets MIL T-28800B, type III class 3, style C specifications, comprised of the following:

Characteristic	Description	
Temperature		
Operating and humidity	-15°C to +55°C/95 (+5, -0) % relative humidity.	
Non-operating	-62°C to +75°C.	
Altitude		
Operating	15,000 feet.	
Non-operating	40,000 feet.	
Humidity (Non-operating)	Five cycles (120 hours) of MIL-Std-810.	
Vibration	Method 507 Procedure 4 (modified).	
Operating	Resonant searches along all three axis at 0.025 inch, frequency varied from 10—55 Hz, 15 minutes. All major resonances must be minimum per axis plus dwell at resonant frequency or 55 Hz for 10 minutes minimum per axis. Instrument secured to vibration platform during test. Total vibration time about 75 minutes.	
Shock (Operating and Non-operating)	Three shocks of 30 G, one-half sine, 11 ms duration, each direction along each major axis. Guillotine-type shocks. Total of 18 shocks.	
Electromagnetic Interference (EMI)	Within limits described in MIL-Std-461.	
	<b>Test Method</b>	<b>Remarks</b>
Conducted emissions	CE01	10 kHz to 20 kHz only.
	CE03 20 kHz to 50 MHz power leads.	Except 30 to 35 kHz, relaxed by 15 dB.
Conducted susceptibility	CS01 30 Hz to 50 kHz power leads.	Full limits.
	CS02 50 kHz to 400 kHz power leads.	Full limits.
	CS06 Spike power leads	Full.
Radiated emissions	RE01 30 Hz to 30 kHz magnetic	Relaxed by 10 dB for fundamental, 2nd, and 3rd harmonic of power line.
	RE02 14 kHz to 10 GHz.	
Radiated susceptibility	RS01 30 Hz to 30 kHz magnetic field.	Full limit.
	RS03 up to 10 GHz.	Full limit.

PHYSICAL CHARACTERISTICS

Characteristics	Description
Weight (standard accessories and cover except manuals)	44 pounds (21.78 kg) maximum.
Dimensions (Fig. 1-1)	
Without front cover and handle or feet.	6.9 X 12.87 X 19.65 inches (17.5 cm X 32.69 cm X 49.91 cm).
With front cover, feet and handle	7.82 X 15.05 X 23.1 inches (handle folded back over instrument), 25.85 inches (handle fully extended).




Fig. 1-1. Dimensions.

ACCESSORIES

Table 1-3  
ACCESSORIES  
STANDARD

Description	Part Number	Qty.	Storage
50 Ω Coax Cable, N to N Conn, 6 Feet	012-0114-00	1	Front Cover
50 Ω Coax Cable, BNC to BNC Conn, 18 Inch	012-0076-00	1	Front Cover
Manual, Operators	070-2726-01	1	
Manual, Operators, Handbook	070-2729-00	1	Front Cover

Table 1-3 (cont)

Description	Part Number	Qty.	Storage
Manual, Service, Volume 1	070-2727-00	1	
Manual, Service, Volume 2	070-2852-00	1	
Adapter, N Male to BNC Female	103-0045-00	1	Front Cover
Fuse 2 A, fast	159-0021-00	1	Front Cover
Fuses 4 A, fast	159-0017-00	2	Front Cover
Power Cord (115 V nominal)	161-0118-00	1	Front Cover
Cord Clamp	343-0170-00	1	Front Cover
Crt Light Filter, Blue	378-0115-03	1	Front Cover
Crt Light Filter, Amber	378-0115-01	1	Front Cover
Crt Light Filter, Grey	378-0115-02	1	Front Cover
Visor Crt	016-0653-00	1	
Crt Mesh Filter	378-0726-01	1	Front Cover
Diplexer Assembly <sup>2</sup>	015-0385-00	1	Front Cover

<sup>2</sup> Option 8 deletes the Diplexer assembly.

## OPTIONAL

<b>General Purpose Waveguide Mixer Set (Option 20)</b>	016-0640-00		Front Cover or Storage Case
Mixer 12.4 to 18 GHz	119-0097-00		
Mixer 18 to 26.5 GHz	119-0098-00		
Mixer 26.5 to 40 GHz	119-0099-00		
50 $\Omega$ Coax Cable, 2 Feet, 3 mm to TNC Connector	012-0748-00		
<b>High Performance Waveguide Mixer Set (Option 21)</b>	016-0662-00		Front Cover or Storage Case
Mixer 18 to 26.5 GHz	016-0631-01		
Mixer 26.5 to 40 GHz	016-0632-01		
50 $\Omega$ Coax Cable, SMA to SMA Connector	012-0649-00		
<b>High Performance Waveguide Mixer Set (Option 22)</b>	016-0657-00		
Mixer 18 to 26.5 GHz	016-0631-01		
Mixer 26.5 to 40 GHz	016-0632-01		
Mixer 40.0 to 60 GHz	016-0634-01		
50 $\Omega$ Coax Cable SMA to SMA Connector	012-0649-00		
<b>Power Cord (220 V nominal)<sup>3</sup></b>	161-0066-01		

<sup>3</sup> See other Power Cord Configurations under Options.

Table 1-3 (cont)

OPTIONAL

Description	Part Number Number	Qty.	Storage
Transit Case (Hard)	016-0658-00		
Transit Case (Soft)	016-0659-00		
Circuit Board Extender Assembly	672-0865-00	1	
Frame (extrusion) for Circuit Board Extender	426-1527-00	1	
Left Extender Board	670-5562-00	1	
Right Extender Board	670-5563-00	2	
Screws, panhead with flat and lockwashers	211-0116-00	6	

## OPTIONS

Options available for the 492 and their resultant changes to the specifications are listed below. Options are factory installed at the time of the initial order. Contact your local Tektronix Field Office for additional information.

### OPTION 1

This option provides calibrated preselection to the first (1st) mixer for the 1.7 to 18 GHz frequency range and limiter protection below 1.8 GHz. Band 1 becomes 100 kHz to 1.8 GHz using an input low-pass filter; the preselector starts at Band 2 (1.7 GHz).

The following changes and additions in electrical characteristics apply:

Table 1-4

### OPTION 1 FREQUENCY CHARACTERISTICS

Frequency Band	Center Frequency Range	LO Frequency & Harmonic (n) <sup>c</sup>	1st IF Response	Frequency Response	Frequency Response Referred to 100 MHz
With Coaxial (direct) Input					
1	0—1.8 GHz <sup>b</sup>	(2072-3872) (1-)	2072 MHz	±2.5 dB (50 kHz—1.8 GHz)  ±1.5 dB (100 kHz—1.8 GHz)	
2	1.7—5.5 GHz	(2529—6329) (1-)	829 MHz	±2.5 dB	±3.5 dB
3	3.0—7.1 GHz	(2171—6271) (1+)	829 MHz	±2.5 dB	±3.5 dB
4	5.4—18.0 GHz	(2072—3872) (3-)	829 MHz	±3.5 dB	±4.5 dB
5	15.0—21.0 GHz	(4309—6309) (3+)	2072 MHz	±5.0 dB	— <sup>a</sup>

Table 1-4 (cont)

## OPTION 1 FREQUENCY CHARACTERISTICS

Frequency Band	Center Frequency Range	LO Frequency & Harmonic (n) <sup>c</sup>	1st IF Response	Frequency Response	Frequency Response Referred to 100 MHz
With External Waveguide Mixers					
6	18.0—26.5 GHz <sup>d</sup>	(2655—4071) (6+)	2072 MHz	±3.0 dB <sup>e</sup>	±6 dB
7	26.5—40.0 GHz <sup>d</sup>	(2443—3793) (10+)	2072 MHz	±3.0 dB <sup>e</sup>	±6 dB
8	40.0—60.0 GHz <sup>d</sup>	3790—5790) (10+)	2072 MHz	±3.0 dB	±6 dB
9	60.0—90.0 GHz <sup>f</sup>	(3861—5862) (15+)	2072 MHz	— <sup>f</sup>	
10	90.0—140 GHz <sup>f</sup>	(3822—6000) (23+)	2072 MHz	— <sup>f</sup>	
11	140—220 GHz <sup>f</sup>	(3728—5890) (37+)	2072 MHz	— <sup>f</sup>	

<sup>a</sup> Frequency range of the coaxial input bands is 100 kHz to 18 GHz; usable with reduced performance, to 50 kHz and 21 GHz.

<sup>b</sup> Frequency range of non-option 1 instruments.

<sup>c</sup> "n" denotes the LO (local oscillator) harmonic number. The + or - sign indicate which side (above or below) the incoming signal is with respect to the effective LO harmonic.

<sup>d</sup> Can be tuned over wider frequency range with unspecified performance.

<sup>e</sup> With TEKTRONIX High Performance Waveguide Mixers.

<sup>f</sup> Dependent on external mixer.

Table 1-5

## OPTION 1 SENSITIVITY

Frequency/Band	Equivalent Input Noise for Resolution Bandwidths			
	1 kHz	10 kHz	100 kHz	1 MHz
100 kHz—7.1 GHz (Bands 1—3)	-110 dBm	-100 dBm	-90 dBm	-80 dBm
5.4—12.0 GHz (Band 4)	-95 dBm	-85 dBm	-75 dBm	-65 dBm
12.0—18.0 GHz (Band 4)	-90 dBm	-80 dBm	-70 dBm	-60 dBm
15.0—21.0 GHz (Band 5)	-85 dBm	—	—	—
18.0—26.5 GHz (Band 6) <sup>a</sup>	-100 dBm	-90 dBm	-80 dBm	-70 dBm
26.5—40.0 GHz (Band 7) <sup>a</sup>	-95 dBm	-85 dBm	-75 dBm	-65 dBm
40.0—60.0 GHz (Band 8) <sup>a</sup>	-95 dBm	-85 dBm	-75 dBm	-65 dBm
60.0—90.0 GHz (Band 9)	External Mixer Dependent			
90.0—140 GHz (Band 10)	External Mixer Dependent			
140—220 GHz (Band 11)	External Mixer Dependent			

<sup>a</sup> High Performance TEKTRONIX Waveguide Mixers.

## ELECTRICAL CHARACTERISTICS OPTION 1 (cont)

Characteristic	Performance Requirement	Supplemental Information
Spurious Responses		
Intermodulation Products		
1.8—18 GHz	$\geq -70$ dBc down from any two on-screen signals within any frequency span.	$\geq -100$ dBc when signals are separated 100 MHz or more.
1.7—1.8 GHz	$\geq -70$ dBc down from any two $-40$ dBm signals within any frequency span.	
Harmonic Distortion (cw signal 1.7—18 GHz)	$-100$ dBc or more for full screen signal (MIN DISTORTION mode).	
LO emission, referenced to input mixer and with zero RF attenuation	$\leq -10$ dBm to 1.7 GHz; and $\leq -70$ dBm from 1.7 to 18 GHz.	
Input Level		
Maximum Safe Input with zero RF attenuation		1 watt or $+30$ dBm.
1 dB Compression Point (minimum):		
1.7—2.0 GHz	$-28$ dBm	
Otherwise	$-10$ dBm	

## OPTION 2

This option provides digital storage. The following are the changes and additions to the instrument:

Multiple memory (A & B) display storage is provided with: Save A, Max Hold, B memory minus Save A memory, digital display averaging, and storage bypass for non-store display.

When digital storage is used, an additional quantization error of 0.5% of full screen must be added to the measured amplitude characteristics (i.e., frequency response, sensitivity, etc.).

## OPTION 3

This option provides first (1st) local oscillator stabilization by phase locking to an internal reference to reduce residual FM when narrow bands are selected. The microcomputer automatically selects phase lock for a span/division of 50 kHz or less in Bands 1 through 3, 100 kHz or less in Band 4, and 200 kHz or less in Bands 5 and above. This option also adds a 100 Hz resolution filter. The instrument characteristics that are changed are listed below.



## ELECTRICAL CHARACTERISTICS OPTION 3 (cont)

Characteristic	Performance Requirement	Supplemental Information			
		Band	Narrow Span	Wide Span	
Frequency Span/Div Range		1—3 (0—7.1 GHz)	500 Hz/Div	200 MHz/Div	
		4—5 (5.4—21 GHz)	500 Hz/Div	500 MHz/Div	
		6 (18—26 GHz)	500 Hz/Div	1 GHz/Div	
		7—8 (26—60 GHz)	500 Hz/Div	2 GHz/Div	
		9 (60—90 GHz)	500 Hz/Div	2 GHz/Div	
		10 (90—140 GHz)	500 Hz/Div	5 GHz/Div	
		11 (140—220 GHz)	500 Hz/Div	10 GHz/Div	
		Two additional positions provide full band display (MAX span) or 0 Hz (time domain display).			
		Accuracy	Within 5% of the span/div selected over the center eight divisions of a ten division display.		
		Resolution	Additional resolution bandwidth of 100 Hz with 7.5:1 shape factor.		
Residual FM (short term) after 2 hour warmup	$\leq (50 \text{ Hz peak-to-peak}) n$ for a period of 20 ms. $n$ is the 1st LO harmonic number used in the 1st mixer conversion, and related to the selected frequency range (band).	No video filter.			
Frequency Drift	25 kHz/hour, fundamental mixing, after 2 hour warmup.				
Sensitivity (100 Hz)	8 dB better than 1 kHz sensitivity.				

## OPTION 8

Deletes External Mixer capability. Standard accessories does not include the Diplexer. Frequency range of the instrument is 50 kHz to 21 GHz.

## OPTION 20

Includes: General Purpose Waveguide Mixers: 12.5 to 40 GHz. Tektronix Part No. 016-0640-00.

Frequency Range	Part No.	Sensitivity: Equivalent Input Noise @ 1 kHz Bandwidth (Typical)
12.4—18 GHz	119-0097-00	-75 dBm
18.0—26.5 GHz	119-0098-00	-70 dBm
25.6—40 GHz	119-0099-00	-60 dBm
Cable: TNC to SMA male connectors, 012-0748-00		
Storage Case: 004-1651-00		

## OPTION 21

Includes: High Performance Waveguide Mixers: 18 to 40 GHz. Tektronix Part No. 016-0662-00.

Frequency Range	Part No.	Sensitivity: Equivalent Input Noise @ 1 kHz Bandwidth (Maximum)	Frequency Response	Referenced to 100 MHz
18.0—26.5 GHz	016-0631-01	-100 dBm	±3.0 dB	±6 dB
26.5—40 GHz	016-0632-01	-95 dBm	±3.0 dB	±6 dB
Cable: SMA to SMA connector, 012-0649-00				
Storage Case: 004-1651-00				

## OPTION 22

Includes: High Performance Waveguide Mixers: 18 to 60 GHz. Tektronix Part No. 016-0657-00


Frequency Range	Part No.	Sensitivity: Equivalent Input Noise @ 1 kHz Bandwidth (Maximum)	Frequency Response	Referenced to 100 MHz
18.0—26.5 GHz	016-0631-01	-100 dBm	±3.0 dB	±6 dB
26.5—40 GHz	016-0632-01	-95 dBm	±3.0 dB	±6 dB
40—60 GHz	016-0634-01	-95 dBm	±3.0 dB	±6 dB
Cable: SMA to SMA male connector, 012-0649-00				
Storage Case: 004-1651-00				

## NOTE

*These characteristics assume that the waveguide mixer is connected to a cw signal source and that the PEAKING control is adjusted for maximum signal amplitude. The signal must be stable (not frequency modulated more than the resolution bandwidth); otherwise, frequency response specifications cannot be met.*

### OPTIONS FOR POWER CORD CONFIGURATION

Tektronix has implemented options that provide international approved power cord and plug configurations. These are shown and illustrated in Fig. 1-2.



2726-11

Fig. 1-2. International power cord and plug configuration for the 492.

# INSTALLATION AND REPACKAGING

## Introduction

This section describes unpacking, installation, power requirements, and repackaging information for the 492 Spectrum Analyzer.

## Unpacking and Initial Inspection

Before unpacking the 492 from its shipping container or carton, inspect for signs of external damage. If the carton is damaged, notify the carrier. The shipping carton contains the basic instrument and its standard accessories. Optional accessories are shipped in separate containers. Refer to the Accessories listing in the Specification section for a complete listing.

If the contents of the shipping container are incomplete, if there is mechanical damage or defect, or if the instrument does not meet operational check requirements, contact your local Tektronix Field Office or representative. If the shipping container is damaged, notify the carrier as well as Tektronix, Inc.

The instrument was inspected both mechanically and electrically before shipment. It should be free of mechanical damage and meet or exceed all electrical specifications. Procedures to check functional or operational performance are in the Operation section. The functional check procedure verifies proper instrument operation. This check should satisfy the requirements for most receiving or incoming inspections. The electrical performance check procedure is part of the Service instructions.

## Preparation for Use

The 492 can be installed in any position that allows air flow in the bottom and out the rear of the instrument. Feet on the four corners allow ample clearance even if the instrument is stacked with other instruments. A fan draws air in through the bottom and expels air out the back. Avoid locating the 492 where paper, plastic, or like material might block the air intake.

The front panel cover for the 492 provides a dust-tight seal. Use the cover to protect the front panel when storing or transporting the instrument. The cover is also used to store accessories and external waveguide mixers. The

cover is removed by first pulling up and in on the two release latches then pulling up on the cover. The door to the accessories compartment is unlatched by pressing the latch to the side and lifting the cover.

The handle of the 492 can be positioned at several angles to serve as a tilt stand, or it can be positioned at the top rear of the instrument between the feet and the rear panel so 492 instruments can be stacked. To position the handle, press in at both pivot points and rotate the handle to the desired position.

### CAUTION

*Removing or replacing the cabinet on the instrument can be hazardous. The cabinet should only be removed by qualified service personnel.*

## Power Source and Power Requirements

The 492 is designed to operate from a single-phase power source that has one of its current-carrying conductors (neutral) at ground (earth) potential. Operating from power sources where both current-carrying conductors are isolated or above ground potential (such as phase-to-phase on a multi-phase system or across the legs of a 110-220 volt single-phase, three-wire system) is not recommended, since only the line conductor has over-current (fuse) protection within the unit. Refer to the Safety Summary at the front of this manual.

The ac power connector is a three-wire polarized plug with the ground (earth) lead connected directly to the instrument frame to provide electrical shock protection. If the unit is connected to any other power source, the unit frame must be connected to an earth ground.

Power and voltage requirements are printed on a back panel plate mounted below the power input jack. The 492 can be operated from either 115 Vac or 230 Vac nominal line voltage with a range of 90 to 132 or 180 to 250 Vac, at 48 to 440 Hz. A multipin (harmonica) type connector on the power supply etched circuit board can be positioned to accommodate either voltage range. When the power supply circuitry is changed to accommodate a different power source, the information plate on the back panel must also be changed to reflect the new power re-

quirements. Refer power input changes to qualified service personnel. Instructions are contained in the Service Information section.

**WARNING**

*Only qualified service personnel should attempt to change the power input requirements. Unfamiliarity with safety procedures can result in personal injury. Refer to the Safety Summary at the front of this manual.*

**Repackaging for Shipment**

When the 492 is to be shipped to a Tektronix Service Center for service or repair, attach a tag showing: owner and address, name of individual at your firm that can be contacted, complete serial number, and a description of the service required. If the original packaging is unfit for use or not available, repackage the equipment as follows:

1. Obtain a carton of corrugated cardboard having inside dimensions that are at least six inches more than the equipment dimensions, to allow for cushioning. Table 2-1 lists instrument weights and carton strength requirements.

2. Install the front cover on the 492 and surround the equipment with polyethylene sheeting to protect the finish.

3. Cushion the equipment on all sides with packing material or urethane foam between the carton and the sides of the equipment.

4. Seal with shipping tape or industrial stapler.

Table 2-1

**SHIPPING CARTON TEST STRENGTH**

Gross Weight		Carton Test Strength	
Pounds	Kilograms	Pounds	Kilograms
0—10	0—3.73	200	74.6
10—30	3.73—11.19	275	102.5
30—120	11.19—44.76	375	140.0
120—140	44.76—52.22	500	186.5
140—160	52.22—59.68	600	223.8

If you have any questions, contact your local Tektronix Field Office or representative.

# OPERATION

This section describes the function of the controls, indicators, and connectors for the 492, operational checkout of the instrument, and a detailed description of some functions.

## CONTROLS, INDICATORS, AND CONNECTORS

Because most operational functions of the 492 are microprocessor controlled, they are switch selected rather than vernier adjusted.

The following describes the function of the controls, indicators, and connectors on the front and rear panels of the 492. Figures 3-1 and 3-2 illustrate their locations. All options are covered in this description and some functions are described in greater depth under Operating Information.

### Front Panel (Fig. 3-1)

- ① **INTENSITY.** Controls brightness of the crt trace and crt readout. (Focus is automatically adjusted electronically.)
- ② **READOUT.** Switches crt readout on or off. All spectrum analyzer parameters are read out except Time/Div. Brightness is proportional to the trace brightness and the ratio can be readjusted by service personnel.
- ③ **GRAT ILLUM.** Switches graticule light on or off.
- ④ **BASELINE CLIP.** When activated, the baseline of the display is clipped or subdued to increase the contrast between the display and the baseline.




Fig. 3-1. Front panel selectors, controls, and connectors.

- 5 **TRIGGERING.** One of four triggering modes can be selected by pushbuttons that illuminate when activated. A SINGLE SWEEP pushbutton plus a READY indicator provide single sweep operation.

**FREE RUN**—When activated, the sweep is free running without regard to trigger signals. When selected, all other triggering modes are canceled.

**INT**—When activated, the sweep is triggered by any signal at the left edge of the display with an amplitude of 1.0 division or more. Other trigger modes are canceled.

**LINE**—When activated, a sample of the ac power line voltage is used to trigger the sweep. All other modes are canceled when selected.

**EXT**—When selected, the sweep is triggered by signals between 0.5 volt peak (minimum) to 50 volts peak (maximum) that are applied through the back panel EXT IN HORIZ/TRIG connector. When EXT is selected, the other modes are canceled.

- 6 **SINGLE SWEEP.** When selected, one sweep is initiated after the sweep circuit has been triggered. Pushing this button does not cancel the trigger modes. The button must be pressed to rearm the sweep circuit after the sweep has run. When single sweep mode is first selected, the present sweep is aborted and the sweep circuit is not armed. To cancel single sweep, press one of the trigger mode pushbuttons.

- 7 **READY.** When SINGLE SWEEP is selected, this indicator lights while the sweep circuit is armed and ready for a trigger signal. The indicator stays lit until the sweep ends.

- 8 **MANUAL SCAN.** When the TIME/DIV selector is in the MNL position, this control will manually scan the spectrum.

- 9 **TIME/DIV.** Selects sweep rates from 5 s/div to 20  $\mu$ s/div in 5-2-1 sequence. This switch also selects AUTO, EXT, and MNL modes.

**AUTO** (automatic)—In this position the sweep rate is selected by the microcomputer to maintain a calibrated display for any FREQ SPAN/DIV, RESOLUTION, and VIDEO FILTER combination.

**EXT** (external input)—This position connects the rear panel EXT IN HORIZ TRIG connector to the horizontal sweep circuit. A voltage ramp of 0 to  $\pm$ 10 volts will sweep 10 divisions of the horizontal (x) axis.

**MNL** (manual)—In this position the horizontal axis can be swept with the MANUAL SCAN control.

- 10 **FREQUENCY.** Tunes the center frequency. Tune rate is proportional to the selected FREQ SPAN/DIV. Any given signal moves across the display at a constant rate for all spans. In MAX span, the tuning range depends on the band; for example, in Band 2 (1.7—5.5 GHz) the frequency dot will not tune to the extreme left edge of the graticule, or in Band 6 (15—21 GHz) the dot will tune only to the right of center.

- 11 **FREQUENCY RANGE** (band). These two pushbuttons shift the center frequency range up or down. Frequency range of the band is displayed on the crt readout.

- 12  **$\Delta$ F.** A convenience for measuring frequency difference between signals (see Operational Procedure). When selected, the frequency readout goes to zero. It will then read out the deviation from this reference as the FREQUENCY is tuned.

- 13 **CAL.** When activated, the frequency readout can be calibrated to center frequency by adjusting the FREQUENCY control for the correct reading. When calibrated, deactivate the CAL mode.

- 14 **DEGAUSS.** When the DEGAUSS button is pressed, current through the tuning coils of the YIG oscillator (1st LO) and YIG preselector (when installed) is reduced to zero to minimize hysteresis effects. This enhances center frequency and display amplitude accuracy. DEGAUSS does not function when the FREQ SPAN/DIV is less than 1 MHz/Div. Degauss the tuning coils after a significant frequency change and before calibrating the center frequency readout.

- 15 **IDENTIFY 500 kHz ONLY.** Signal identify feature is functional only when the FREQ SPAN/DIV is 500 kHz. When activated (button lit) true signals will change amplitude each sweep; images and spurious response signals will shift horizontally

or off screen. To ensure that the signal is changing amplitude every sweep, decrease the sweep rate so each sweep can be analyzed; or, if the instrument has digital storage, activate SAVE A, VIEW B.

When the true signal is centered under the dot marker after degauss, the FREQUENCY readout is the signal frequency (within specification). Degauss by pressing DEGAUSS at FREQ SPAN/DIV setting of 1 MHz or 2 MHz.

- 16 **PHASE LOCK (Option 3).** The 1st LO is locked to a stable internal reference and the 2nd LO swept to reduce residual FM in narrow spans; the button lights when phase lock is active. In narrow spans phase lock can be turned off or back on by pressing the button. Spans for which the microcomputer automatically selects phase lock are:

Band	Span/Div
1, 2, 3	50 kHz and below
4	100 kHz and below
5 and above	200 kHz and below

- 17 **AUTO RESOLUTION.** This is a pushbutton that activates automatic bandwidth selection for the selected FREQ SPAN/DIV and TIME/DIV and VIDEO FILTER. An internal microcomputer selects bandwidth to maintain a calibrated display. When the TIME/DIV is in AUTO mode, resolution bandwidth becomes a function of the FREQ SPAN/DIV selection.

- 18 **FREQUENCY SPAN/DIV.** This is a continuous detented control that selects frequency span/div. Span/div is indicated by the crt readout. Range of the span/div selection depends on the frequency band and options. Table 3-1 lists the range for the various bands and options. Selection is a 5-10-20 sequence plus MAX span and 0 Hz span positions.

When MAX span is selected, the span displays the full band. Sweep beyond the band is clamped to the baseline. A dot marker near the top of the screen indicates the position on the span of the crt frequency readout. This dot and frequency point will be center screen when the FREQ SPAN/DIV is reduced below MAX span position. When zero span is selected Time/Div is read out instead of Span/Div.

Table 3-1

**SPAN/DIV RANGES VERSUS BAND AND OPTION**

Band	Narrow Span/Div		Wide Span/Div
	Standard	Option 3	All Instruments
1—3 (0—7.1 GHz)	5 kHz	500 Hz	200 MHz
4—5 (5.4—21 GHz)	10 kHz	500 Hz	500 MHz
6 (18—26 GHz)	20 kHz	500 Hz	1 GHz
7—8 (26—60 GHz)	50 kHz	500 Hz	2 GHz
9 (60—90 GHz)	100 kHz	500 Hz	2 GHz
10 (90—140 GHz)	100 kHz	500 Hz	5 GHz
11 (140—220 GHz)	100 kHz	500 Hz	10 GHz

- 19 **RESOLUTION BANDWIDTH.** This is also a continuous detented control that selects resolution bandwidth. Bandwidth is indicated by crt readout. Range of selection is 1 kHz to 1 MHz in decade steps. An additional resolution band-

width of 100 Hz, when Option 3 is installed, is provided. Changing the resolution bandwidth with this control deactivates AUTO RESOLUTION.



20 **VERTICAL DISPLAY.** These four pushbuttons select the display mode. The crt readout indicates scale factor

**10 dB/DIV**—When activated, the dynamic range of the display is a calibrated 80 dB with each major graticule representing 10 dB.

**2 dB/DIV**—Increases resolution so that each major graticule division represents 2 dB.

**LIN**—Selects a linear display between zero volts (bottom graticule line) and the reference level (top graticule line) scaled in volts/division. See **REFERENCE LEVEL**.

**PULSE STRETCHER**—Increases the fall time of pulse signals so very narrow pulses in a line spectrum display can be seen. The effect is most apparent for discrete signals analyzed at resolution bandwidths that are narrow compared to the span; **PULSE STRETCHER** may be necessary for digital storage of such signals, especially if they are averaged.

21 **VIDEO FILTER.** One of two (**NARROW** and **WIDE**) filters can be activated to reduce the video bandwidth and reduce high frequency components for display noise averaging. The **NARROW** filter is approximately 1/300th of the selected resolution bandwidth; the **WIDE** filter is about 1/30th the bandwidth. Activating either filter cancels or deactivates the other filter. Press the pushbutton to switch the filters off.

22 **DIGITAL STORAGE (Option 2).** Five pushbuttons and one control operate the digital storage functions. With none of the pushbuttons activated, the 492 display is not stored.

**VIEW A, VIEW B**—When either or both of these pushbuttons are selected, the pushbutton illuminates and the contents of memory A and/or memory B are displayed. With **Save A** mode off, all memory locations are displayed and updated continuously. Data in A memory is interfaced with data from B memory.

**SAVE A**—When activated, this mode holds data in A memory and inhibits further updating. With **SAVE A** and **VIEW A** active, data in A memory is displayed but not updated, serving as a reference to compare contents of B memory.

**B—SAVE A**—When activated, the differential (arithmetic difference) of data in B memory and the saved data in memory A is displayed. **SAVE A** mode is activated and **SAVE A** button illuminated. The zero difference point is nominally set at the middle graticule line with positive differences displayed above this line and negative differences below. (The zero difference position on screen is internally switch selectable.)

**MAX HOLD**—When activated the digital storage memory retains the maximum signal amplitude at each memory location. This permits visual monitoring of signal frequency and amplitude at each memory location over an indefinite period of time. This feature is used to measure drift, stability, and record peak amplitudes.

**PEAK/AVERAGE**—This control selects the amplitude at which the vertical display is either peak detected or averaged. Video signals above the level set by the control (shown by a horizontal line or cursor), are peak detected and stored; video signals below the cursor are digitally averaged and stored. See **Peak/ Average Control (Digital Storage)** under **General Information**.

23 **MIN RF ATTEN dB and REFERENCE LEVEL.** These two concentric controls select the minimum RF attenuation (in dB) between the RF INPUT and the first mixer and the display reference level (top of graticule). Both reference level and actual RF attenuation (which may exceed the **MIN RF ATTEN dB** setting) are displayed by the crt readout.

**REFERENCE LEVEL** is a continuous control that requests the microcomputer to change the reference level one step each time the control is moved one detent. The normal steps are 10 dB in the 10 dB/DIV vertical display mode and 1 dB in the 2 dB/DIV mode. When the vertical display is changed to **LIN**, the reference level remains the same. It now changes in 1 dB steps and the scale factor is 1/8 the voltage equivalent of the reference level. See **FINE** for the fine resolution steps.

**MIN RF ATTEN dB** selects the lowest value of attenuation allowed; actual RF attenuation is set by the microcomputer according to the algorithm selected by the **MIN NOISE/MIN DISTORTION** button. If RF attenuation is increased by changing **MIN RF ATTEN**, the microcomputer automatically changes **IF** gain to maintain the current reference level.

- 24 **UNCAL.** This indicator lights when the display amplitude is no longer calibrated (e.g., selecting a sweep rate that is not compatible with the frequency span/div and resolution bandwidth).
- 25 **LOG and AMPL CAL.** These adjustments calibrate the dynamic range of the display. The LOG calibrates the logarithmic gain in dB/Div, the AMPL calibrates the reference level of the top graticule line at the top of the screen.
- 26 **FINE.** A pushbutton used to select fine incremental steps for the REFERENCE LEVEL. The reference level does not change when FINE is switched off even if it is not at a normal step for the display mode.


Vertical Display Mode	FINE Increment
10 dB/DIV	1 dB
2 dB/DIV	0.25 dB ( $\Delta A$ mode) <sup>a</sup>
LIN	Voltage equivalent to 1 dB

<sup>a</sup> For  $\Delta A$  mode description see Delta A Mode under General Operating Information.

- 27 **MIN NOISE/MIN DISTORTION.** One of two algorithms is selected to control attenuator and IF gain. MIN NOISE (button illuminated) reduces the noise level by reducing attenuation 10 dB and reducing IF gain 10 dB. MIN DISTORTION reduces IM distortion due to input mixer overload. To observe any change, the RF ATTEN, displayed by the crt readout, must be 10 dB higher than that set by the MIN RF ATTEN selector.

In MIN DISTORTION mode (button not illuminated) distortion is minimum.

- 28 **POWER.** Pull type switch that switches the main power supply on.


- 29  **RF INPUT.** A 50  $\Omega$  coaxial input connector for signals 21 GHz or below. The maximum, non-destructive input signal level to the input mixer is +13 dBm or 30 mW. Signals above -10 dBm may cause signal compression.

**CAUTION**

The maximum rating of the RF attenuator is +30 dBm (1 watt average, 75 watts peak, pulse width

21  $\mu s$ , with a duty cycle that does not exceed 0.001). Burn-out occurs above 1 watt.

If MIN NOISE is activated and RF ATTEN is 60 dB, the +30 dBm rating could be exceeded. If the input signal level is increased for a full screen display the input level will be +40 dBm. Reduce high level signals with external attenuators if the RF attenuator rating may be exceeded. Use external attenuators and the MIN RF ATTEN to reduce the level into the 1st mixer to -10 dBm or less. Input voltage to the input mixer must not contain any dc component. Refer to Signal Application (under General Operating Instructions) discussed later in this section.

- 30 **POSITION** . These controls position the display on the horizontal and vertical axes.

- 31 **CAL OUT** (Calibrator output). The source of a calibrated -20 dBm ( $\pm 0.3$  dB) 100 MHz ( $\pm 0.01\%$ ) signal, and a comb of frequency markers 100 MHz apart. The calibrated 100 MHz marker is used as a reference for calibrating reference level and log scale. The comb of 100 MHz markers is used to check span and frequency readout accuracy.

- 32 **OUTPUT 1st and 2nd LO.** These connectors provide access to the output of the respective local oscillators. The connectors must be terminated into 50  $\Omega$  when they are not connected to some external device.

- 33 **EXT MIXER.** When the EXT MIXER button is activated, bias is provided out the EXT MIXER port for external waveguide mixers. The IF output from the external mixer is then applied through the EXT MIXER port to the 2nd converter. External mixer connection and operation is described under General Operating Information.

**CAUTION**

Do not exceed mixer input limits. Refer to the external mixer operating instructions at the end of the General Operating Information part of this section.

- 34 **PEAKING.** This control varies mixer bias for external mixers in the EXT MIXER mode. If the 492 has a preselector (Option 1), the control also adjusts the preselector filter tracking for the 1.7 to 21 GHz frequency range (Bands 2-5). In both cases it is adjusted for maximum signal amplitude. Refer to External Mixer Operation for more detailed instructions.

Rear Panel (Fig. 3-2)

- ① **EXT IN HORIZ/TRIG.** Dc coupled input for horizontal drive voltages and ac coupled for trigger signal. A 0 to -10 volt ramp produces full sweep. 0.5 to 50 volt peak signals are required for trigger (0.1  $\mu$ s minimum pulse width), 15 Hz to 1 MHz. Selection as to HORIZ or TRIG mode depends on front panel TRIGGERING and TIME/DIV selections.
- ② **PROBE POWER.** Provides operating voltage ( $\pm 15$  V and +5 V) for active probes, such as TEKTRONIX P6056 or P6021. Refer to probe instruction manuals for operating procedures.
- ③ **HORIZ (output).** Source of a signal that is 0.5 V for each division of display.
- ④ **VERT (output).** Source of a signal that is 0.5 V for each division of display.
- ⑤ **PEN LIFT.** TTL compatible, nominal +5 V provided to lift the pen of a chart recorder.
- ⑥ **10 MHz IF (+20 dBm max).** Access to the 10 MHz IF signal. Output level is about -10 dBm with a full screen signal at -30 dBm reference level, maximum output is +20 dBm.
- ⑦ **J104 ACCESSORY.** Possible future applications for the 492 may use this connector.




Fig. 3-2. Rear panel connectors.

2726-13

## TURN ON PROCEDURE AND PREPARATION FOR USE

The following procedure initiates a display and calibrates center frequency readout, display reference level, and dynamic range.

### 1. Initial Turn On

a. Connect the 492 power cord to an appropriate power source (see Power Requirements under Installation instructions) and switch POWER on. Allow three to four minutes for the instrument to warm up and stabilize before proceeding. Note that the crt readout is functioning (see Fig. 3-3).

When POWER is switched on (power up), the operating functions and modes of the 492 initialize to the following "power up" state.

Vertical Display	10 dB/DIV
FREQUENCY	0.00 MHz
REF LEVEL	+30 dBm
RF ATTENUATION	60 dB
FREQUENCY RANGE	0.0—4.2 GHz (0.0—1.8 GHz with Option 1)
AUTO RESOLUTION	On
RESOLUTION BAND- WIDTH	1 MHz
FREQ SPAN/DIV	MAX
TRIGGERING	FREE RUN
READOUT	On
Digital Storage (Option 2)	VIEW A/VIEW B on
All other pushbuttons	Inactive or off




Fig. 3-3. CRT readout for power-up state.

b. Set MIN RF ATTEN to 0 dB and PEAK/AVERAGE control fully down. Set the TIME DIV to AUTO, the REFERENCE LEVEL to -20 dBm, and adjust the INTENSITY for a display with the desired brightness. Note, the RF ATTEN readout is now 10 dB.

c. Apply the CAL OUT to the RF INPUT by connecting a 50  $\Omega$  coaxial cable between the CAL OUT connector and the RF INPUT.

d. In the MAX frequency span mode, a dot marker in the upper portion of the screen indicates the location on the display to which the 492 center frequency is tuned. With a frequency readout of 0 MHz, it will be in the upper left portion of the screen. Adjust the center FREQUENCY control and note the dot marker move across the display.

e. Note the comb of 100 MHz markers at the left side of the display (see Fig. 3-4). Tune the dot marker to a position above the first 100 MHz marker.

f. Change the FREQ SPAN/DIV to 100 MHz. Note that the dot marker is now centered horizontally and the 100 MHz signal is at or near center screen.

g. Position the dot marker to the graticule centerline with the horizontal (◀▶) position control. Adjust the center frequency control to tune the signal over the dot marker.

h. Press the 2 dB/DIV Vertical Display button, then position the baseline of the display to the bottom graticule line with the vertical (▲) POSITION control.




Fig. 3-4. Typical display of calibrator markers in MAX SPAN position.

## 2. Calibrate Center Frequency Readout

a. Change the Vertical Display to 10 dB/DIV. If the instrument has phase lock, ensure that the PHASE LOCK is deactivated or the FREQ SPAN/DIV is above 50 kHz (recommend setting of 10 MHz).

b. Tune the FREQUENCY to place the 100 MHz calibrator signal over the center-span dot marker. Reduce FREQ SPAN/DIV to 1 MHz, press the DEGAUSS button, and fine tune center frequency.

### NOTE

*Degauss function is inoperative when the FREQ SPAN/DIV is less than 1 MHz.*

c. Press the CAL button to activate the calibration function (button illuminated); then adjust the FREQUENCY tuning control for a readout of 100 MHz. Press the CAL button to deactivate the CAL mode and increase FREQ SPAN/DIV to 20 MHz.

d. Check—center frequency accuracy at other multiples of the 100 MHz calibrator signal. Approach each check point from the low frequency side and degauss the tuning coils of the oscillator as each check point is approached. Readout should be within  $\pm(5 \text{ MHz} + 20\% \text{ span/div})$  or  $\pm(0.2\% \text{ of center frequency} + 20\% \text{ of span/div})$  whichever is greater.

e. Return the frequency to 100 MHz.

## 3. Calibrate Reference Level and Dynamic Range

a. With the 100 MHz calibrator signal tuned to center screen and the REF LEVEL at  $-20 \text{ dBm}$ , set the FREQ SPAN/DIV to 20 kHz, as indicated by the crt readout.

b. Alternately switch Vertical Display from 10 dB/DIV to 2 dB/DIV and adjust AMPL CAL so the peak amplitude of the signal is the same for each logarithmic display mode.

c. With the Vertical Display mode at 10 dB/DIV, set the top of the calibrator signal to the top graticule line with the LOG CAL adjustment so the signal level indicates  $-20 \text{ dBm}$ .

d. Check—the display log scale over 50 dB of dynamic range by switching REF LEVEL in 10 dB steps from  $-20 \text{ dBm}$  to  $+30 \text{ dBm}$  and noting that the display amplitude decreases 10 dB or one division per step.

e. Set the REF LEVEL at  $-20 \text{ dBm}$  and the Vertical Display to 2 dB/DIV.

f. With the REF LEVEL selection in coarse mode, change the REF LEVEL to  $-10 \text{ dBm}$  in 1 dB steps. Check that the amplitude and REF LEVEL readout reduces 1 dB/step for a total of  $10 \text{ dB} \pm 1.3 \text{ dB}$ .

g. Return the REF LEVEL to  $-20 \text{ dBm}$  and activate FINE for the  $\Delta A$  mode.

h. Change the REF LEVEL 2.00 dB. Check that the REF LEVEL readout changes in 0.25 dB steps and the display amplitude decreases  $2 \text{ dB} \pm 0.5 \text{ dB}$  (1 division  $\pm 0.25$  division) to match the 2 dB change in readout. Return the REF LEVEL to 0.00 dB.

i. Change the display mode to 10 dB/DIV, switch the REF LEVEL from  $-20 \text{ dBm}$  to  $-10 \text{ dBm}$ . Check that the REF LEVEL changes in 1 dB increments and the total change in display amplitude is  $10 \text{ dB} \pm 1.3 \text{ dB}$  (1 division  $\pm 0.13$  division).

j. Return the REF LEVEL to  $-20 \text{ dBm}$  and cancel FINE for coarse steps.

## 4. Check Span Accuracy and Linearity

Span accuracy is the displacement error of calibrator markers from the center reference over  $\pm 4$  divisions of span. Linearity is the displacement error of markers from their specified points over the display area, with the 1st graticule line as the reference.

a. Set the FREQ SPAN/DIV to 50 MHz and the frequency to about 500 MHz. Tune the 500 MHz calibrator marker to center screen so it is below the center frequency dot marker.

b. Check span accuracy by noting that the 100-MHz markers are within 5% of their reference graticule line over the center eight divisions.

c. Tune one of the markers to the 1st graticule line. Check linearity by noting that the displacement between successive markers, with respect to the FREQ SPAN/DIV setting over the display area, does not exceed 5%.

## FUNCTIONAL OR OPERATIONAL CHECK

This procedure uses minimum test equipment to check instrument operating modes, functions, and basic performance. The procedure serves to check for instrument malfunctions and should satisfy most incoming inspection or pre-operational check-out requirements. A detailed Performance Check, which requires extensive test equipment, is part of the Calibration Procedure in the service instructions. This operational check also familiarizes the user with instrument operation.

### Equipment Required

The internal calibrator for the 492 is an accurate amplitude and frequency reference source. It, along with a very accurate internal attenuator, is used as the reference source for checks in this procedure. The following fixtures, which are part of the standard accessories, are also required.

1. Adapter: N male to BNC female.
2. 50  $\Omega$  coaxial cable: 18 inch, BNC to BNC connectors.

### Preliminary Preparation

a. Perform the initial calibration described under Turn On Procedure.

b. Set the front panel controls as follows:

BASELINE CLIPper	Off
TRIGGERING	FREE RUN
TIME/DIV	AUTO
FREQUENCY RANGE	0.0—4.2 GHz (0.0—1.8 GHz Option 1)
FREQUENCY	100 MHz
VERTICAL DISPLAY	10 dB/DIV
FREQ SPAN/DIV	1 MHz
AUTO RESOLUTION	On
MIN RF ATTEN	0 dB
MIN NOISE (pushbutton)	Off (MIN DISTORTION)
FINE (pushbutton)	Coarse (not illuminated)
REF LEVEL	-20 dBm
DIGITAL STORAGE (Option 2)	
VIEW A	On
VIEW B	On
MAX HOLD	Off
SAVE A	Off
B—SAVE A	Off
PEAK/AVERAGE	Fully ccw

c. Allow the instrument to warm up for at least 30 minutes before proceeding with this check.

### 1. Check Operation of Front Panel Pushbuttons and Controls

The following procedure checks functions activated by front panel pushbuttons and that the buttons illuminate when the function is active. Operation of the front panel controls is also checked.

With the CAL OUT signal applied to the RF INPUT, tune the 100 MHz, -20 dBm signal to center screen. Reduce the FREQ SPAN/DIV to 100 kHz keeping the signal centered on screen. Press or change the following pushbuttons and controls and note their effect.

**INTENSITY.** Rotate the control through its range and note crt beam brightness change.

**READOUT.** Inactive state, no crt readout. Active state, crt readout of REF LEVEL, FREQUENCY, FREQ SPAN/DIV, VERT DISPLAY, RF ATTEN, FREQ RANGE, and RESOLUTION BANDWIDTH. The INTENSITY control changes brightness.

**GRAT ILLUM.** Inactive state, no graticule lights. Active state, graticule lighted.

**BASELINE CLIP.** Inactive, no clipping of the display baseline. Active, display intensity at the baseline is clipped (subdued).

**TRIGGERING.** Triggering mode is activated by pressing one of four pushbuttons. Pressing any one of the buttons cancels or deactivates the other mode.

**FREE RUN.** Active, trace free runs.

**INT.** Active, trace displayed when signal or noise level at left edge is  $\geq 1.0$  division.

**LINE.** Active, trace triggered at power line frequency.

**EXT.** Active, trace runs only when an external signal  $\geq 0.5$  volt peak is applied to the back panel EXT IN connector.

**SINGLE SWEEP.** Pressing this button to activate single sweep aborts the current sweep, pressing the button again arms the sweep generator and lights READY, which remains lighted until the sweep completes. The analyzer makes a single sweep of the selected spectrum when the conditions determined by TRIGGERING are met. Single sweep mode is canceled when any TRIGGERING button is pressed. The effect of SINGLE SWEEP may be more apparent if VIEW A, VIEW B, and B—SAVE A are off.

**TIME/DIV.** Selects sweep rate and manual scan operation. In MNL position, MANUAL SCAN control should vary the crt beam across the full horizontal axis of the crt graticule.

**VERTICAL DISPLAY.** Display modes are activated by three pushbuttons. Pressing any of these buttons cancels the other mode.

**10 dB/DIV.** Active, display is a calibrated 10 dB/division, 80 dB dynamic range. Calibration is checked later in this procedure.

**2 dB/DIV.** Active, display is calibrated 2 dB/division, 16 dB dynamic range. Calibration is checked later in this procedure.

**LIN.** Active, display is linear between the reference level (top of graticule) and zero volt (bottom of graticule); the crt VERT DISPLAY reads out in volts/division.

**PULSE STRETCHER.** Active, increases the fall time of video signals to make narrow pulses on the display easier to see. With FREQ SPAN/DIV at MAX, TIME/DIV at 5 ms and Digital Storage off, the markers should increase in brightness when PULSE STRETCHER is active.

**VIDEO FILTER.** Two filters, independently selected to provide WIDE (1/30th) or NARROW (1/300th) of the resolution bandwidth for noise reduction.

**DIGITAL STORAGE (Option 2).** Either or both sections of memory can be selected to provide digital storage. When either or both are activated, signal amplitude should remain constant. Vary the PEAK/AVERAGE control and note that noise level below the PEAK/AVERAGE cursor is averaged.

**VIEW A.** Active, half of digital storage memory is displayed.

**VIEW B.** Active, the B section or other half of memory is displayed.

When both VIEW A and VIEW B are active, contents of A and B memory are interlaced and displayed. Both sections are updated each sweep. Update of A memory depends on the state of SAVE A.

**SAVE A.** Active, contents in A memory are saved and not updated. Verify operation by changing REF LEVEL and observe that the VIEW A display does not change when VIEW B is inactive.

**MAX HOLD.** Active, stores maximum signal amplitude at each memory location. Verify operation by changing FREQUENCY or REF LEVEL and note that the maximum level at each location is retained.

**B—SAVE A.** Active, the difference between updated data in B section of memory and that saved in A is displayed. Verify by saving data in A, then changing the reference level and pressing B—SAVE A; only the difference can be observed by canceling VIEW A and VIEW B. The reference (zero difference) level is normally set at graticule center, but can be internally adjusted by service personnel.

**PEAK/AVERAGE.** When digital storage is activated with VIEW A or VIEW B, this control positions a horizontal line or cursor on the display. Signals above the cursor are peak detected; signals below the cursor are averaged. The cursor should position anywhere within the graticule window.

**IDENTIFY 500 kHz/ONLY.** When activated, the FREQUENCY SPAN must be 500 kHz/div. True signals change amplitude, spurious signals shift horizontally. Set FREQ SPAN/DIV at 500 kHz and press the button. Note the action of the calibrator marker. True signals should not shift horizontally.

**PHASE LOCK (Option 3).** Active to reduce residual FM when narrow spans are selected. The button lights when active; pressing the button turns phase lock off. The microcomputer automatically selects phase lock for a span/division of 50 kHz or below in bands 1 through 3, 100 kHz or below in band 4, and 200 kHz or below in bands 5 and above.

**AUTO RESOLUTION.** When activated, RESOLUTION BANDWIDTH changes so bandwidth is compatible with FREQ SPAN/DIV selection. Check by changing FREQ SPAN/DIV and noting that RESOLUTION BANDWIDTH

changes. **UNCAL** indicator should not light over the SPAN/DIV range if TIME/DIV selector is in AUTO position.

**FREQUENCY SPAN/DIV.** As this control is rotated clockwise or counterclockwise, frequency span/div should change from 0 to MAX in 5-2-1 sequence. Display should indicate this change.

**RESOLUTION BANDWIDTH.** As this control is rotated, resolution bandwidth should range from 1 MHz to 1 kHz (100 Hz Option 3) in decade steps.

**$\Delta$ F.** When activated, center frequency readout initializes to 0 MHz. The frequency difference, to a desired signal or point on the display, can now be determined by tuning that point to center screen and noting the readout. Check by measuring the difference between calibrator markers. Deactivate and note that the readout returns to center frequency. If the frequency is tuned below "0", the readout will indicate (-) sign.

**DEGAUSS.** When pressed, hysteresis in the local oscillator tuning system is reduced. Switch **FREQ SPAN/DIV** to 1 MHz and tune the calibrator marker to center screen. Note the signal position, then press the **DEGAUSS** button. The signal should shift horizontally and then return to a new location. Press again and the signal should return to the same new location. Return **FREQ SPAN/DIV** to 100 MHz.

**FREQUENCY RANGE.** Two pushbuttons that shift the 492 frequency bands. Press the  $\blacktriangle$  button and note the up shift of bands; then press the  $\blacktriangledown$  button and note that the bands shift down to the 0 to 4.2 GHz range (0 to 1.8 GHz Option 1).

**CAL.** Checked when performing Turn On Procedure.

**REFERENCE LEVEL.** Continuous control that requests the microcomputer to change the reference level one step for each detent. In the 10 dB/DIV Vertical Display mode, the steps are 10 dB. When **FINE** is activated, the steps are 1 dB. In the 2 dB/DIV mode, the steps are 1 dB or 0.25 dB for the **FINE** mode. When **FINE** is activated in the 2 dB/DIV mode, the  $\Delta$ A mode is operational. The **REFERENCE LEVEL** goes to 0.00 dB then steps in 0.25 dB increments from an initial 0.00 dB reference level.

Set the **MIN RF ATTEN** to 0 dB, Vertical Display to 10 dB/DIV, and rotate the **REFERENCE LEVEL** control counterclockwise to +30 dBm then clockwise to -120 dBm. Note the change in the display. Return the **REF**

**LEVEL** to -20 dBm and note that 10 dB of **RF ATTEN** is switched in at -20 dBm.

**MIN RF ATTEN.** Sets the minimum amount of RF attenuation. Changing **REF LEVEL** will not decrease RF attenuation below that set by the **MIN RF ATTEN** selector.

**FINE.** When activated, **REF LEVEL** switches in 1 dB increments for 10 dB/DIV display mode, and 0.25 dB for 2 dB/DIV display mode. In the 2 dB/DIV display mode, **FINE** actuates  $\Delta$ A mode. See Delta A Mode description under General Operating Information.

**MIN NOISE/MIN DISTORTION.** One of two algorithms is selected to control attenuator and IF gain. **MIN NOISE** (button illuminated) reduces the noise level by reducing attenuation 10 dB and decreasing IF gain 10 dB. **MIN DISTORTION** reduces IM distortion due to input mixer overload. To observe any change, the **RF ATTEN**, displayed by the crt readout, must be 10 dB higher than that set by the **MIN RF ATTEN** selector.

**UNCAL.** This light comes on when the display is uncalibrated. Set the **TIME/DIV** to 50 ms, deactivate the **AUTO RESOLUTION**, and set the **RESOLUTION BANDWIDTH** to 10 kHz. **UNCAL** should light and remain lit until the **FREQ SPAN/DIV** is reduced to 200 kHz or the **RESOLUTION BANDWIDTH** is increased to 1 MHz. Return the **TIME/DIV** to **AUTO** and activate the **AUTO RESOLUTION**. Set the **FREQ SPAN/DIV** to 100 MHz.

**EXTERNAL MIXER/PEAKING.** In active mode, bias for external waveguide mixers is provided at the **EXT MIXER** connection. Activate the External Mixer mode by changing the **FREQUENCY RANGE** to 18—26 GHz and then measure the bias with a VOM between the center conductor and ground of the **EXT MIXER** port. Bias should range from about -1.5 to +0.5 volts as the **PEAKING** control is varied.

If the instrument has a preselector (Option 1), the control also varies the preselector tuning to augment tracking for the coaxial bands (0—21 GHz).

This completes the functional check of the front panel controls and pushbuttons.

## 2. Check Frequency Readout Accuracy

Readout accuracy  $\pm(5 \text{ MHz} + 20\% \text{ of span/division})$  or  $\pm(0.2\% \text{ of center frequency} + 20\% \text{ of span/division})$ , whichever is greater.



**NOTE**

*Due to hysteresis in the 1st (YIG) oscillator, accuracy of the frequency readout should be checked after the tuning coil has been degaussed by pressing the DEGAUSS button when the FREQ SPAN/DIV is either 2 MHz or 1 MHz, before reducing the FREQ SPAN/DIV to narrower span settings.*

a. Calibrate the frequency readout as described under Initial Operation. Now check frequency readout accuracy by tuning to successive 100 MHz calibrator markers. As a given marker is tuned toward center screen, degauss the oscillator tuning coils by pressing the DEGAUSS button with the FREQ SPAN/DIV either at 2 MHz or 1 MHz. Reduce the FREQ SPAN/DIV to 500 kHz and center the marker under the frequency dot before noting the readout for center frequency. Readout should be within  $\pm(5 \text{ MHz} + 20\%$  of span/division) of the calibrator marker; for example, the 10th marker should represent 1 GHz. In Bands 2 and 3 it may be necessary to activate WIDE Video Filter and MIN NOISE. Set the FREQ SPAN/DIV to 10 MHz and use RESOLUTION BANDWIDTH of 100 kHz. Adjust PEAKING, if the 492 has Option 1 (preselector), as you approach each check point. In some instruments the calibrator harmonic signals may be very small. Either ignore these check points or try reducing the resolution bandwidth more to increase sensitivity. Sweep time must be very long when using narrow resolution bandwidths.

b. Since the other bands operate on harmonics of the oscillator fundamental, accuracy or error will be the same as that measured for the fundamental band multiplied by the harmonic number (n) of the band or frequency range in use.

**3. Check Frequency Span/Div Range and Accuracy**

Range is in a 5-10-20 sequence and depends on frequency band and option as shown in Table 3-2.

Span accuracy specifies the maximum displacement error of calibrator markers from the center screen reference over the center 8 divisions of span. Linearity is the displacement error between successive markers with respect to the FREQ SPAN/DIV setting over the center 8 divisions of span.

a. Tune the center FREQUENCY to the 1.0 GHz (900 MHz for Option 1 instruments) calibrator harmonic signal and set the REF LEVEL at -30 dBm. Set the FREQ SPAN/DIV to 200 MHz and activate AUTO RESOLUTION.

b. Check—for 2 GHz (1.8 GHz Option 1) of span. Check span/div accuracy by noting marker displacement from the respective graticule line. Deviation should not exceed 5% or 1 minor division.

c. Check—the range and accuracy for FREQ SPAN/DIV selections to 10 MHz using the calibrator 100 MHz comb of markers as the reference.

**Table 3-2**

**NARROW AND WIDE SPANS VERSUS FREQUENCY BAND**

Band	Narrow Span		Wide Span
	Standard	Option 3	All Instruments
1-3	10 kHz/Div	500 Hz/Div	200 MHz/Div
4-5	50 kHz/Div	500 Hz/Div	500 MHz/Div
6	50 kHz/Div	500 Hz/Div	1 GHz/Div
7-8	100 kHz/Div	500 Hz/Div	2 GHz/Div
9	200 kHz/Div	500 Hz/Div	2 GHz/Div
10	500 kHz/Div	500 Hz/Div	5 GHz/Div
11	500 kHz/Div	500 Hz/Div	10 GHz/Div

Accuracy: Within 5% of selected span/div over center 8 divisions of display.

d. Change FREQUENCY RANGE and check range for all bands as per specification. Accuracy will remain consistent with that noted in part b of this step.

**4. Check Resolution Bandwidth and Shape Factor (bandwidth, 1 kHz to 1 MHz ±20% in decade steps; shape factor, 7.5:1 or less)**

a. Tune the center frequency to 100 MHz and center the 100 MHz calibrator signal. Set the REF LEVEL to -20 dBm. FREQ SPAN/DIV to 500 kHz, RESOLUTION BANDWIDTH to 1 MHz, TIME/DIV to AUTO, Vertical Display mode to 2 dB/DIV, and MIN NOISE on.

b. Measure the 6 dB bandwidth (see Fig. 3-5A). Bandwidth should equal 1 MHz ±200 kHz.

c. Change the Vertical Display mode to 10 dB/DIV.

d. Measure the -60 dB bandwidth (see Fig. 3-5B). Calculate the shape factor by dividing the -60 dB bandwidth by the -6 dB bandwidth. The shape factor should equal 7.5:1 or less.

e. Switch RESOLUTION BANDWIDTH to 100 kHz and the FREQ SPAN/DIV to 100 kHz.

f. Check the bandwidth and shape factor of the 100 kHz filter by repeating the foregoing procedure.

g. Switch the RESOLUTION BANDWIDTH to the remaining selections, adjusting the FREQ SPAN/DIV as necessary to check the bandwidth and shape factor of each selection. Bandwidth should be within 20% of that selected and the shape factor 7.5:1 or better.

**5. Check Reference Level Gain and RF Attenuator Steps**

**NOTE**

*This procedure checks functional operation but does not verify accuracy. If deviation seems significant, refer to qualified service personnel for a performance check.*

a. With the calibrator signal applied to the RF INPUT, set the front panel controls as follows:

FREQUENCY RANGE	Band 1 (0—4.2 or 0—1.8 Option 3)
FREQ SPAN/DIV	10 MHz




Fig. 3-5. Display to illustrate how bandwidth and shape factors are determined.

RESOLUTION BANDWIDTH	1 MHz
Video Filter	NARROW
Vertical Display	10 dB/DIV
REF LEVEL	-30 dBm
MIN RF ATTEN	0 dB
TIME/DIV	AUTO
MIN NOISE	ON

b. Tune FREQUENCY to the 200 MHz marker, then decrease FREQ SPAN/DIV and RESOLUTION BANDWIDTH to 100 kHz, keeping the marker centered.

c. Check the attenuator by increasing REF LEVEL to 40 dBm in 10 dB steps, noting that the signal peak drops 1 division per step and RF ATTEN increases to 60 dB in 10 dB steps.

d. Return REF LEVEL to -20 dBm and increase MIN RF ATTEN dB to 60 dB, noting that the noise level rises about 1 division per step.

e. Check IF gain steps by switching REF LEVEL from -20 dBm to +40 dBm; observe that the signal peak and noise level decreases 1 division per step.

f. Activate FINE; then check that the trace rises in 1 dB (0.1 division) steps as REF LEVEL is changed to +30 dBm.

g. Switch FINE off and reduce MIN RF ATTEN dB to 0. Deactivate MIN NOISE and note that the noise level rises about 1 division as the RF ATTEN changes 10 dB (50 dB to 60 dB) when MIN NOISE is switched off (REF LEVEL should not change and the trace peak should remain relatively constant).

h. Adjust REF LEVEL to position the signal peak near the top of the graticule. Change Vertical Display mode to 2 dB/DIV.

i. Check that REF LEVEL and signal amplitude changes in 1 dB steps as the REFERENCE LEVEL control is rotated.

j. Turn on FINE and note that the REF LEVEL readout changes to \*0.00 dB. This denotes the delta A mode. Check that REF LEVEL and signal amplitude changes in 0.25 dB steps as the REFERENCE LEVEL is rotated.

k. Switch FINE off and set REF LEVEL to a multiple of 10 dBm. Change Vertical Display mode to 10 dB/DIV.

6. Check Sensitivity (refer to Table 3-4 or 3-5)

NOTE

*Sensitivity is specified according to the input mixer average noise level. The 492 calibrator signal is the reference used to calibrate the display. Accuracy of the calibrator output level can be verified using a 100 MHz bandpass filter with known loss and an accurate power meter.*

a. Set the front panel controls as follows:

FREQUENCY RANGE	0.0—4.2 GHz (0.0—1.8 GHz, Option 1)
FREQUENCY	Within Band 1
Vertical Display	10 dB/DIV
MIN RF ATTEN dB	0
REF LEVEL	-30 dBm
FREQ SPAN/DIV	5 MHz
RESOLUTION	
BANDWIDTH	1 MHz
TIME/DIV	0.5 s
Digital Storage (Option 2)	VIEW A/VIEW B
MAX HOLD	Off
PEAK/AVERAGE Cursor	Fully cw
Video Filter	WIDE

b. Disconnect the calibrator signal from the RF INPUT.

c. Check—noise level below the -30 dBm reference level. Noise floor (level) should not be above that specified in Table 3-3 or 3-4.

d. Change REF LEVEL to -40 dBm and FREQ SPAN/DIV to 1 MHz.

e. Check—noise level of 10 kHz and 100 kHz resolution bandwidths. Compare these levels with those listed in Table 3-3 or 3-4.

f. Change the REF LEVEL to -60 dBm, FREQ SPAN/DIV to 10 kHz, VIDEO FILTER to NARROW, and TIME/DIV to AUTO.

Table 3-3

SENSITIVITY WITHOUT PRESELECTOR

Average Noise Level dBm (max)					
Frequency Range	Resolution Bandwidth				Option 3
	1 MHz	100 kHz	10 kHz	1 kHz	100 Hz
100 kHz—4.2 GHz (Bands 1—3)	-85	-95	-105	-115	-123
5.4—18 GHz (Band 4)	-70	-80	-90	-100	-108
15—21.0 GHz (Band 5)	-65	-75	-85	-95	-100
18—26 GHz (Band 6) <sup>a</sup>	-70	-80	-90	-100	-110
26—40 GHz (Band 7) <sup>a</sup>	-65	-75	-85	-95	-105
40—60 GHz (Band 8) <sup>a</sup>	-65	-75	-85	-95	-105
60—90 GHz (Band 9)	Function of external mixer				
90—140 GHz (Band 10)	Function of external mixer				
140—220 GHz (Band 11)	Function of external mixer				

<sup>a</sup> High performance TEKTRONIX waveguide mixers.

g. Check—average noise level to 1 kHz resolution bandwidth against that listed in Table 3-3 or 3-4.

h. If Option 3 is installed, change RESOLUTION BANDWIDTH to 100 Hz and FREQ SPAN/DIV to 500 Hz.

i. Check—sensitivity for Option 3 instruments. See Table 3-3 or 3-4.

j. Return PEAK/AVERAGE control to the fully ccw position and switch VIDEO FILTER off.

NOTE

*This procedure may be used to check sensitivity characteristics with optional external waveguide mixers when an accurate signal source is used to establish the reference.*

Table 3-4  
SENSITIVITY OPTION 1

Frequency Range	Average Noise Level dBm (max)				Option 3
	Resolution Bandwidth				
	1 MHz	100 kHz	10 kHz	1 kHz	100 Hz
100 kHz—7.1 GHz (Bands 1—3)	-80	-90	-100	-110	-120
5.4—18.0 GHz (Band 4)	-65	-75	-85	-95	-105
15.0—21.0 GHz (Band 5)	-	-	-	-85	-100
18.0—26.5 GHz (Band 6)*	-70	-80	-90	-100	-110
26.5—40.0 GHz (Band 7)*	-65	-75	-85	-95	-105
40.0—60.0 GHz (Band 8)*	-65	-75	-85	-95	-105
60.0—90.0 GHz (Band 9)	Function of external mixer				
90.0—140 GHz (Band 10)	Function of external mixer				
140—220 GHz (Band 11)	Function of external mixer				

\* High performance TEKTRONIX Waveguide Mixers.

**7. Frequency Drift or Stability** (within 200 kHz/hour without phase lock, within 25 kHz/hour with Option 3 phase lock)

**NOTE**

*This measurement and residual FM are dependent on oscillator stability. Therefore, the instrument must have at least a two hour warm-up period.*

a. Select a Vertical Display of 10 dB/DIV, FREQUENCY RANGE 0.0—4.2 GHz (0—1.8 GHz, Option 1), REF LEVEL of -20 dBm, FREQ SPAN/DIV of 10 MHz, TIME/DIV at AUTO, and activate AUTO RESOLUTION.

b. Connect the CAL OUT signal to the RF INPUT then tune one of the calibrator comb markers to center screen. Press the DEGAUSS button to remove any residual magnetism as the signal is centered.

c. Set REF LEVEL for a signal amplitude of about seven divisions. Reduce FREQ SPAN/DIV to 100 kHz, cancel PHASE LOCK (if Option 3 instrument), and adjust FREQUENCY to center the signal on screen.

d. If the instrument has digital storage (Option 2), select VIEW B, SAVE A, and MAX HOLD; cancel VIEW A.

e. Check frequency drift over the specified period. If the instrument has digital storage, activate MAX HOLD and measure drift as the increased width of the marker over the specified time period. (See Fig. 3-6.) If the instrument does not have digital storage, measure drift by observing the change in position of the marker over the time period.




Fig. 3-6. Typical display using Digital Storage with MAX HOLD activated to measure drift.

f. If the instrument has phase lock (Option 3), activate PHASE LOCK and reduce FREQ SPAN/DIV to 10 kHz. Recenter the marker signal on screen. Cancel MAX HOLD, then check drift as in parts d and e.

g. If the instrument has digital storage, cancel MAX HOLD and SAVE A.

8. Check Residual FM (within 1 kHz for 20 milliseconds without phase lock, and within 50 Hz for 200 milliseconds with Option 3 phase lock)

a. Set the FREQUENCY RANGE to Band 1 (0—7.1 GHz or 0—4.1 GHz Option 1). If the 492 has phase lock, cancel PHASE LOCK and center the calibrator signal with the FREQUENCY control. (Increase FREQ SPAN/DIV to locate signal if off screen, then return to 10 kHz/div.)

b. Set RESOLUTION BANDWIDTH to 10 kHz and the Vertical Display to LIN. Activate FINE and adjust REF LEVEL for a full screen display.

c. Position the marker signal with the FREQUENCY control so the slope (horizontal versus vertical excursion) of the response can be measured as illustrated in Fig. 3-7A. SINGLE SWEEP may be advantageous to freeze the display if the instrument has digital storage.

d. Switch FREQ SPAN/DIV to zero (time domain), TIME/DIV to 20 ms, and adjust FREQUENCY to position the display near center screen as shown in Fig. 3-7B. Note the peak-to-peak amplitude of the display within any horizontal division, scaling the vertical deflections according to the slope estimated in part c. Residual FM must not exceed 1 kHz for 20 ms.

e. If the instrument has phase lock, proceed as follows:

1) Switch TIME/DIV to AUTO and activate PHASE LOCK. Increase FREQ SPAN/DIV to bring the signal on screen, then reduce the span to 500 Hz/div and the RESOLUTION BANDWIDTH to 1 kHz. Keep the signal centered with the FREQUENCY control.

2) Calculate the slope as described in part c.

3) Switch to zero span and TIME/DIV to 200 ms/div; then measure residual FM using the same technique described above. Residual FM, for Option 3 instruments, must not exceed 50 Hz for a 200 ms period.




Fig. 3-7. Display that illustrates how to measure residual (incidental) FM.

## 9. Digital Storage (Option 2)

a. Set the front panel controls as follows:

REF LEVEL	-10 dBm
Vertical Display	10 dB/DIV
RESOLUTION BANDWIDTH	1 MHz
FREQUENCY	100 MHz
TIME/DIV	AUTO
Digital Storage (Option 2)	VIEW A

b. With the calibrator signal applied to the RF INPUT, tune the signal to center screen while reducing the FREQ SPAN/DIV to 200 kHz. Change the Vertical Display to 2 dB/DIV, then activate SAVE A.

c. Change the REF LEVEL to +30 dBm and activate VIEW B. Display B of the calibrator signal should be 2 dB less than display A.

d. Activate B-SAVE A and check display. Display should show the difference between display B and display A (see Fig. 3-8).




Fig. 3-8. Using Digital Storage to measure differential between two displays.

e. Deactivate both SAVE A and B-SAVE A functions and activate MAX HOLD.

f. Change both FREQUENCY and REF LEVEL and note that MAX HOLD function retains and holds the maximum signal amplitude and frequency excursion.

g. Deactivate MAX HOLD and select VIEW A and AUTO BANDWIDTH. Reduce the FREQ SPAN/DIV to 100 kHz and change the Vertical Display to 10 dB/DIV.

h. Vary the PEAK/AVERAGE control to shift the cursor over the screen height and note that signal and noise are averaged below the cursor.

This completes the operational check of the 492.

## GENERAL OPERATING INFORMATION

### Crt Light Filters

Three light filters are supplied with the 492 accessories: amber, grey, and blue. Filter selection depends on am-

bient light conditions, light reflections, and operator's viewing needs. The filter is installed by pulling the top of the plastic mask out and placing the filter behind the crt bezel. It is best to remove the light filter when taking display photographs.

### Intensity Level, Focus, and Beam Alignment

Operate the instrument with the intensity level no higher than that required to clearly observe the display. Trace alignment and beam focus are internal adjustments that must be performed by qualified service personnel.

Intensity level for some displays, such as pulsed spectra, must be higher which may produce a display with a very bright baseline. This bright baseline can be clipped or subdued by activating the BASELINE CLIP.

### Signal Application

Signal frequencies up to 21 GHz are applied through a short, high quality, 50 Ω coaxial cable, to the RF INPUT connector. These signals pass through an internal RF attenuator to the first (1st) mixer. Option 1 version of the 492 has a filter selector that automatically selects either a low-pass filter or tuned preselector (depending on frequency band) between the RF attenuator and the 1st mixer.

Signals above 21 GHz are applied to an external waveguide mixer. The output of the waveguide mixer is then applied through the EXT MIXER port to the second converter of the 492. Waveguide mixers and their application are described later.

**RF INPUT Connector.** The nominal input impedance of the coaxial RF INPUT is 50 Ω. Because cable losses can be significant at microwave frequencies, it is important to keep the cables as short as possible. Impedance mismatch between the signal source and the RF INPUT will produce reflections and degrade flatness, frequency response, sensitivity, and increase spurious responses. Impedance mismatch can be caused by poor connections, incorrect signal source impedance, long or low quality coaxial cable, etc. When optimum flatness or frequency response is desired and signal strength is adequate, set the MIN RF ATTENUATION for 10 dB or more. The addition of the attenuator helps minimize reflections to improve the input characteristics.

### CAUTION

With MIN NOISE activated and 60 dB of MIN RF ATTEN, the REF LEVEL can be set to +40 dBm. The front end of the 492 is specified at +30 dBm maximum. Do not increase input signal level to full

screen with a REF LEVEL of  $-40$  dBm because this will exceed the attenuator rating. DC input is limited to zero (0) volt.

As stated in the preceding CAUTION, too much power can cause signal compression or if excessive, it can destroy the 1st mixer. Signals greater than  $-30$  dBm or  $-20$  dBm in MIN NOISE mode, should be attenuated by the RF ATTENUATOR. Signals above the safe input level ( $+30$  dBm) must be attenuated by external attenuators. Ensure that the frequency range of external attenuators is adequate.

Signal levels of  $-10$  dBm or more ( $-28$  dBm Option 1, for 1.7 to 1.8 GHz range, see Specifications) may be compressed. This may degrade signal reference level measurements and generate spurious responses.

Spurious responses can be minimized if the signal amplitude is kept within the graticule window. A recommended procedure is to select a REF LEVEL that places stronger signals just within the graticule window. In some cases, as previously described, it may be best to add RF ATTENUATION.

**Amplitude Conversion.** The 492 reads out signal levels in dBm. A conversion chart shown in Fig. 3-9 provides a way to determine input signal levels from a voltage or power source.

**Connecting to 75  $\Omega$  Source:** The 492 can be used with a 75  $\Omega$  signal source at the lower frequencies (100 kHz—1 GHz) by using a 75  $\Omega$  to 50  $\Omega$  minimum loss attenuator. This attenuator is available as an optional accessory (refer to the optional accessories list in the catalog or Accessory page in the manual for ordering information).




Fig. 3-9. Volts-dBm-watts conversion chart for 50  $\Omega$  impedance.

Sensitivity and power levels are often rated in dBm (dB with reference to 1 mW regardless of impedance). Sensitivity and power levels for 75  $\Omega$  systems are usually rated in dBmV (dB with reference to 1 mV across 75  $\Omega$ ). Figure 3-10 is a circuit diagram of a suitable matching pad for this purpose. Figure 3-11 shows the relationship between 50  $\Omega$  and 75  $\Omega$  units, with matching attenuators included; the conversion is described as follows:

- (1) dBmV (75  $\Omega$ ) = dBm (50  $\Omega$ ) + 54.47 dB.  
For example,  $-60$  dBm (50  $\Omega$ ) + 54.47 dB =  $-5.5$  dBmV (75  $\Omega$ ).
- (2) dBm (75  $\Omega$ ) = dBm (50  $\Omega$ ) + 5.72 dB.  
For example,  $-60$  dBm (50  $\Omega$ ) + 5.72 dB =  $-54.3$  dBm (75  $\Omega$ ).

(3) For some applications you may wish to know the relationship between dBm and dB $\mu$ V.

For 50  $\Omega$  systems dB $\mu$ V = dBm + 107 dB.

### Resolution Bandwidth, Frequency Span, and Sweep Time

Resolution is the ability of a spectrum analyzer to display discrete frequency components within a frequency span. This ability is a function of the analyzer bandwidth, sweep time, frequency span, and incidental FM. Frequency span and sweep time are normally selected to provide the minimum resolution bandwidth



setting for a particular cw signal. Bandwidth also has an effect on noise level. As the bandwidth decreases, signal-to-noise ratio or sensitivity increases. Maximum sensitivity is therefore attained at the narrow resolution bandwidth settings.

As the analyzer sweep rate is increased, a critical rate is reached where both sensitivity and resolution are degraded. Therefore, sweep time for a calibrated display is dependent on resolution bandwidth and the frequency span.

In spans other than MAX SPAN, frequency span is symmetrical about the center frequency. In MAX SPAN the display represents the full frequency range of the selected band. A frequency dot above the display indicates the location on the spectrum of the FREQUENCY readout. The frequency span used depends on the application. Wide spans are normally used to monitor a frequency spectrum for spurious signals, check harmonic content, etc. Narrow spans are used to identify the characteristics around a particular signal, such as modulation side bands, bandwidth, power line related distortion, etc. When wide spans are used, sweep rate on non-store displays is usually increased to eliminate flicker. This requires wider resolution bandwidths. Narrow spans, used to observe signal phenomena, usually call for narrow resolution bandwidths and therefore slow sweep speeds.

The 492 features microcomputer circuitry that selects sweep rate and resolution bandwidth to correlate with the selected frequency span. When both TIME/DIV and RESOLUTION are in the AUTO mode, the display is calibrated for each FREQ SPAN/DIV selection. The AUTO position of the TIME/DIV selector ties the sweep speed to the analyzer span/div and resolution bandwidth. The AUTO mode of the RESOLUTION BANDWIDTH optimizes bandwidth for the selected FREQ SPAN/DIV and TIME/DIV settings unless either is outside the range of correction. When this occurs, the UNCAL indicator lights and a ">" symbol prefixes the REF Level readout on the crt display.

When analyzing pulse signals, a wider bandwidth than that provided by AUTO is usually desired. The resolution bandwidth should be on the order of 1/10 the side lobe frequency width, or the reciprocal of the pulse width, in order to ensure adequate resolution. The RESOLUTION BANDWIDTH is usually set for optimum main lobe detail after the sweep rate has been selected.




Fig. 3-10. Circuit of a 75 Ω to 50 Ω matching pad (AC coupled).




Fig. 3-11. Graph to illustrate relationship between dBm, dBmV, and dBμV (matching attenuator included where necessary).

### Using the PEAKING Control

The PEAKING control adjusts bias for the EXT MIXER port and the preselector tracking for the instruments with Option 1. It is adjusted for maximum conversion or maximum signal amplitude. This control has a marked effect on performance when operating in the higher frequency ranges. Mixer peaking, when Option 1 is installed, must be adjusted before relative amplitude and sensitivity measurements are made when operating above Band 1 (1.8 GHz). Frequency response and flatness are also affected; therefore, after any significant frequency

change, it is good practice to degauss, then adjust PEAKING for maximum signal amplitude. Degauss with FREQ SPAN/DIV of 2 MHz or 1 MHz.

### Using the Signal Identifier

Conversion in the 1st mixer generates many spurious responses. This is due to the multiple harmonics of the local oscillator converting input signals to an intermediate frequency within the bandpass of the IF. This is especially true for the basic 492 (without the Option 1 preselector) and when the waveguide mixers are used.

To help identify true signals, the 492 features an 'identify' mode. With the FREQ SPAN/DIV at 500 kHz, press the IDENTIFY 500 kHz/ONLY button. True signals will alternately shift vertically while spurious signals shift horizontally or off screen. Identify mode will operate only when the frequency span is 500 kHz/div.

### Using the Video Filters

The video filter restricts the video bandwidth so that noise or beat signals are reduced (see Fig. 3-12). When signals are closely spaced, the filter may be useful to reduce modulation between two signals so they can be more easily analyzed. The filters can also be used to average the envelope of pulsed RF spectra that has a relatively high prf (pulse repetition frequency); however, because the filter is basically an integrating circuit, selecting a Video Filter when measuring low prf spectra produces poor results.

The WIDE filter reduces the bandwidth to approximately 1/30th the selected resolution bandwidth; the NARROW filter about 1/300th. Using the filter may require a reduction in the sweep rate to maintain a calibrated display. Again the UNCAL indicator will light if the sweep speed is too fast for video filtering.

### Time Domain Operation

When the FREQ SPAN/DIV is reduced to zero, the analyzer functions as a tunable receiver to display time domain characteristics within the capabilities of the resolution bandwidth. The TIME/DIV selector can now be used to analyze such characteristics as modulation pattern, pulse repetition rates, etc.

### Triggering the Display

Triggering is usually FREE RUN for spectrum displays; however, it may be desirable or necessary to trigger the display when the event is time related to some source or when the frequency span has been reduced to zero for time domain analysis. In the FREE RUN mode the sweep will not synchronize with any input signal.




Fig. 3-12. Integrating the display with the Video Filter.

The sweep can be triggered internally from the vertical or video signal, at the line frequency rate of the power supply, or from an external signal applied to the EXT IN HORIZ/TRIG jack on the back panel. The amplitude of trigger signal required to trigger the sweep is one (1.0) division or more, for internal triggering, and 0.5 volt to a maximum of 50 volts (dc + peak ac) for external triggering.

Trigger source is selected by activating one of the triggering pushbuttons. In addition to the four trigger source selections, SINGLE SWEEP mode can be selected. The sweep will run once after the circuit has been armed and trigger signal arrives. The READY indicator lights when the circuit is armed and waiting for a trigger signal and remains lit until the sweep has run. Pushing the

SINGLE SWEEP button once activates single sweep mode; pushing it again arms the trigger circuit so it is ready for a trigger signal.

### Sweeping the Display

Horizontal sweep for the display is either internal or from an external sweep source. Sweep rate and source are selected by the TIME/DIV switch. When the TIME/DIV switch is in the AUTO position, the sweep rate is controlled by an internal microcomputer.

When the TIME/DIV is in the EXT position, a signal source of 0 to +10 volts, applied to the EXT IN HORIZ/TRIG connector, will sweep the crt beam the full 10 division graticule span. The input is dc coupled, sensitivity is 1 V/div. External input impedance is about 10 k $\Omega$ .

The beam can be positioned by the MANUAL SCAN control when the TIME/DIV is in the MNL position (see Manual Scan of the Spectrum that follows).

### Manual Scan of the Spectrum

Manual scan is used to examine a particular point or portion of a display such as one of the null points of a frequency modulation spectrum or where a slow sweep of the full span would take unnecessarily long. When the TIME/DIV control is set in the MNL position, the display may be swept with the MANUAL SCAN control. The sweep scan is usually first calibrated in one of the timed sweep positions. Note that with a wide span/div and/or a narrow resolution bandwidth setting, it is possible to scan too rapidly to achieve an accurate display. Also, digital storage can give unpredictable results when used with the MNL SCAN mode. Digital storage is updated only when scanning toward the right.

### Reference Level, RF Attenuation, and Vertical Display

A change in the REFERENCE LEVEL control requests the microcomputer to change the display reference level—the absolute amplitude represented by the top of the crt graticule. The microcomputer selects the gain distribution (IF gain and input RF attenuation) for the new reference level according to the setting of the FINE, VERTICAL DISPLAY mode, MIN RF ATTEN dB, and MIN NOISE/MIN DISTORTION selectors.

The amount of attenuation between the RF INPUT and the first mixer, set by the microcomputer, is based on the reference level requested and the mode of the MIN RF ATTEN dB and MIN NOISE/MIN DISTORTION selectors. The microcomputer assumes the MIN RF ATTEN dB

selection is the minimum attenuation required for the expected signal levels. It does not reduce RF attenuation below this value. It also selects the best ratio of RF attenuation and IF gain according to the MIN NOISE/MIN DISTORTION mode (see description that follows). MIN RF ATTEN selects the lower limit reference level range. As MIN RF ATTEN dB is increased, the lower limit reference level is raised an equal amount. At 0 dB minimum attenuation, the lower limit reference level is -123 dBm. At 10 dBm minimum attenuation, the reference level goes to -113 dBm, etc.

The reference level increments depend on the Vertical Display mode and FINE selector mode. Reference level steps for the log displays are 10 dB and 1 dB with FINE off, and 1 dB and 0.25 dB with FINE activated (0.25 dB steps apply to the  $\Delta A$  mode). For LIN displays with FINE off, the microcomputer selects the reference level, which is the equivalent of an 8-division signal, where the bottom of the crt graticule is zero volts and the top of the crt graticule is eight times the vertical display factor. The display factor changes in a 1-2-5 volts/division sequence. For LIN displays with FINE on, the reference level changes in 1 dB steps and the scale factor is 1/8 the voltage equivalent of the reference level.

### Delta A Mode

To select this mode, activate 2 dB/DIV and FINE; the REF LEVEL readout becomes \*0.00 dB and the REFERENCE LEVEL steps in 0.25 dB increments.

The  $\Delta A$  mode is useful for measuring relative amplitude differences of signals more accurately. This is because the gain distribution (IF gain and RF attenuation) is not changed when  $\Delta A$  mode is activated. The REF LEVEL is changed by shifting the log amplifier offset. The measurement range of the  $\Delta A$  mode is at least from 10 dB above to 40 dB below the reference level established when the mode was activated; however, the overall instrument display characteristic of -123 dBm to +30 dBm cannot be exceeded. The asterisk in the REF LEVEL readout remains until the  $\Delta A$  mode gain distribution is changed.

The  $\Delta A$  mode is canceled when either FINE or 2 dB/DIV are deactivated, or a selector that could change gain distribution (MIN RF ATTEN or MIN NOISE) is changed. The analyzer also deactivates  $\Delta A$  mode when EXT MIXER or an external mixer frequency range is selected.

Signals with large differences in amplitude that are within the  $\Delta A$  range can be compared without the distortion usually introduced when signals are driven off-screen. Signals shifted off-screen by changes in the  $\Delta A$  reference level are not overdriving the input because the

attenuator and IF gain are not changed, thus the mixers do not see any change in signal levels due to the  $\Delta A$  reference level changes.

To measure amplitude level differences of two signals:

- 1) Select  $\Delta A$  mode by activating 2 dB/DIV and FINE.
- 2) Using the REF LEVEL control, set the larger amplitude signal to a graticule line.
- 3) Press the FINE pushbutton twice to deactivate and re-activate the  $\Delta A$  mode.
- 4) Using the REF LEVEL control, set the lower amplitude signal to the same graticule line established in step 2.
- 5) The REF LEVEL readout displays the dB difference in amplitude level.

### MIN NOISE/MIN DISTORTION

This pushbutton selects one of two algorithms that control attenuator and IF gain settings. MIN NOISE minimizes noise level while MIN DISTORTION minimizes input mixer overload. To observe any change when MIN NOISE is activated, the RF ATTEN crt readout must be 10 dB higher than that set by the MIN RF ATTEN selector.

#### CAUTION

*With MIN NOISE activated and 60 dB of MIN RF ATTEN, the REF LEVEL can be set to +40 dBm. The front end of the 492 is specified at +30 dBm maximum. Do not increase input signal level to full screen with a REF LEVEL of +40 dBm because this will exceed the attenuator rating. DC input is limited to zero (0) volt.*

### Digital Storage (Option 2)

Digital storage provides a smooth (flicker free) display. Two complete events can be stored. One of these can be saved and then compared to subsequent updated information. A MAX HOLD feature updates the stored data in memory when the new input is of higher amplitude thus allowing monitoring and graphic plotting of display changes with time. Vertical information can be divided by a cursor, or horizontal line, that is positioned with the PEAK/AVERAGE control. Above the cursor, video information is peak detected and displayed; below the cursor, signal averaging occurs. The average (number of samples) is a function of sweep speed. The slower the sweep, the greater the number of samples averaged. This feature suppresses noise in that portion below the cursor and allows full peak detection of vertical data above the cursor. An intensified spot on the cursor indicates the horizontal position at which memory is being updated.

When digital storage is used, an additional quantization error of 0.5% of full screen must be added to the amplitude performance characteristics (i.e., frequency response, sensitivity, etc.).

Digital storage memory is functionally divided into two sections—A and B. Data can be stored in A or B or in both. There are 512 horizontal locations in A and 512 horizontal locations in B. When both are displayed, the origin of B is shifted such that the A and B coordinates are interlaced to provide 1024 display increments. Data in memory is continually updated with each sweep so the display, when viewing A or B, is always current.

When SAVE A function is activated, data in A memory is held in storage and only B memory is updated. This inhibition takes place whether A is displayed or not. This mode captures an event or waveform for comparison with a subsequent event displayed by VIEW B mode. In this mode all of A memory is displayed, then all of B, each by a separate sweep.

When B—SAVE A is activated, the contents of data in B memory minus the contents saved in A are displayed. This provides the comparison of the two events by presenting the algebraic difference of the two displays. This convenient mode can be used to align filters or other devices when tuning for a null. The reference waveform is stored in A and the unknown in B. If the device under test is active, the B waveform may be larger than the reference which results in a shift in the zero reference line. The position of the zero reference can be selected with an 8-bit digital switch. The reference level is normally set mid-screen so positive and negative quantities can be observed. Qualified service personnel can position the reference anywhere within the graticule window.

MAX HOLD causes the digital memory to be updated only if the new input is of higher magnitude than the former (B memory only if SAVE A is active). This allows monitoring of signals that may change with time and provides a graphic record of amplitude/frequency excursions.

Signal averaging is useful for suppressing noise. The number of samples averaged per digitized slot (increment) is a function of the spectrum analyzer sweep rate. The slower the sweep speed, the more samples averaged per resolution bandwidth. Resolution bandwidth also affects the amplitude difference between peak detected and average levels of cw signals. When the resolution bandwidth is less than 1/30th the span/division (e.g., 100 kHz or less with 5 MHz span/div) there will be significant difference between peak and average amplitude levels of cw signals. The peak value will be the true value, the average value will be in error, especially if only A or B is displayed. It is best to run digital storage with both A and B interlaced when using narrow resolution bandwidth with wide frequency spans.

To analyze signal amplitude level, set the cursor at least 1/4 division below the signal peak. To average noise, set the cursor at least 1/4 division above the noise level.

## Waveguide Mixers and External Diplexer

### Introduction

Two types of waveguide mixers are available: lower cost, general purpose mixers that cover the microwave bands, and TEKTRONIX High Performance Waveguide Mixers that cover both microwave and millimeter-wave frequency bands. The 18 to 26.5 GHz and 26.5 to 40 GHz frequency ranges are considered microwave bands; above 40 GHz are the millimeter-wave bands. Improved frequency response and sensitivity characteristics for the High Performance Waveguide Mixers (see Specification Section 1) makes them a better choice for critical measurements in the microwave bands, whereas the General Purpose Waveguide Mixers can be used for initial measurements or as backup units for the high performance mixers.

Typical broadband frequency response for the High Performance Waveguide Mixers is  $\pm 3$  dB to 60 GHz. The mixers are optimized for flatness over each waveguide band.

Both the High Performance and General Purpose microwave mixers have field replaceable diodes. The millimeter-wave mixers are not field repairable and must be returned to Tektronix, Inc. for repair.

Waveguide mixers are optional accessories. The external diplexer is part of the standard accessories (see Options in Section 1). Each mixer is supplied with a data sheet; mixer sets (i.e., Option 20, 21, or 22) include an instruction manual (070-0206-01) that contains detailed description and operational procedure.

### Handling



To prevent damage to the waveguide mixers, read the following instructions fully and carefully before installing or using the mixers.

Handle the waveguide mixers with care. The mixer diode is sensitive to static discharges and excessive rf energy. The maximum input level to all waveguide mixers is +10 dBm or 10 mW cw rf (see Specifications for pulse characteristics). Bias polarity for Tektronix waveguide mixers is negative-going. Check bias requirements of non-Tektronix mixers before connecting them to the 492 Spectrum Analyzer.

Ensure that the shorting cap is installed when the mixer is not in use and install the flange cover on the mixer before returning it to the storage box. The mixer diode can also be destroyed by mechanical vibration or shock.

Do not use an ohmmeter to test or check the mixer diode. The voltage across the test leads of many ohmmeters is capable of destroying the diode.

Use care to avoid scratching the flange surface; these can degrade the performance.

### Installation

The waveguide mixer is connected in the rf system of the analyzer as shown in Fig. 3-13. Physically the mixer is bolted to a waveguide flange at or near the rf signal source. A flexible cable is used to connect the mixer to the external Diplexer. Since the capacitance of the cable is capable of storing enough energy to destroy the mixer diode, connect the cable first to the EXT MIXER port of the Diplexer to discharge it before connecting to the mixer. For best performance use the recommended cable; do not extend its length.

When installing the mixer, make sure the flange surfaces are clean and free of scratches. Be sure to install and tighten all flange screws. This care will minimize input vswr and provide optimum frequency response.




Fig. 3-13. External (Waveguide) mixer installation.

Never apply more than  $+10$  dBm of continuous rf energy to the input of the waveguide mixer port (see Specifications). The waveguide mixers saturate at  $-20$  dBm (typical), therefore, little is gained with inputs above this level. If the input level is unknown, use the general purpose mixer or appropriate waveguide attenuator and RF power meter to test the input level.

The mixers require  $-7$  dBm (min) to  $+15$  dBm (max), typically  $+10$  dBm, of LO signal with a variable bias from  $-2.0$  V to  $+0.5$  V through a current limiting resistor, to meet sensitivity and frequency response characteristics.

### CAUTION

*The Tektronix mixers are designed to use with Tektronix spectrum analyzers. The mixer diode may be damaged if the unit is connected to an analyzer with a different bias supply.*

Activate the EXT MIXER pushbutton and adjust PEAKING for maximum response amplitude. The PEAKING control adjustment will produce more than one maximum through its range. Adjust for the maximum peak.

### CAUTION

*When EXT MIXER is activated or the FREQUENCY RANGE is in the waveguide bands (18 GHz and above), the internal RF attenuator is bypassed. The MIN RF ATTEN control is inactive and the REF LEVEL cannot be set to a value less than  $-30$  dBm.*

## Operation

When using the waveguide mixers, many spurious responses will be displayed. A typical MAX SPAN display, generated by a  $-30$  dBm signal at 40 GHz, is shown in Fig. 3-14. The true response must be identified before any analyses can be made.

If the approximate frequency of the signal is known, select the FREQUENCY RANGE (band) and adjust FREQUENCY for the correct readout; then open the display with the SPAN/DIV selector and adjust FREQUENCY so the desired signal is at center screen. Adjust PEAKING for maximum response over its range; then reduce SPAN/DIV to 500 kHz so IDENTIFIER feature can be used to verify that the signal is a true response.

When the IDENTIFIER mode is active, the display will alternately sweep between two levels (approximately two divisions as shown in Fig. 3-15). If the displayed signal is a

true response, there will be no horizontal offset between the two displays. If the signal is a spurious response, there will be a significant horizontal offset between the two displays (off screen in some cases). As the harmonic conversion number  $n$  increases, for the higher bands, the offset may be small. The accuracy of the center frequency readout is also a function of this conversion number; therefore, a cavity (resonant) wavemeter is recommended to aid in determining frequency. This is especially apparent in the millimeter-wave bands where the conversion number ( $n$ ) is 10 or higher.

When the frequency of the input signal is unknown, the use of a cavity wavemeter is highly recommended; otherwise, each signal must be analyzed with the IDENTIFIER until the true response is found. The wavemeter dip, when it is tuned to the input frequency, can be easily seen on a power meter connected into the signal path; otherwise, use the 2 dB/DIV display mode and adjust the wavemeter for a dip indication of the responses.

The PEAKING control serves to adjust the tracking of the preselector (Option 1 instruments) for the coax bands and to adjust external mixer bias for the External Mixers in the waveguide bands. Adjust the PEAKING for maximum response through its range. (There are usually several peaks over the range of adjustment; adjust for the maximum peak.) After any change in input signal frequency, readjust PEAKING to maintain sensitivity and frequency response characteristics.




Fig. 3-14. Typical display generated by a signal into the waveguide mixer.




Fig. 3-15. Identifier mode displays.

### Analyzing Signals

The following are operational precautions to observe when analyzing displays.

1. If the analyzer has digital storage, ensure that the cursor is at the bottom of the display or screen for peak detection. Digital storage averaging can obscure signals in the noise on wide spans.

2. Use AUTO RESOLUTION with care when measuring absolute amplitude level. Always use a bandwidth wider than the incidental FM level of the signal source.

3. The reference level is calibrated to compensate for the nominal conversion loss of the waveguide mixers in each waveguide band. Slight variations between mixers result in an amplitude accuracy of about  $\pm 6$  dB, including the  $\pm 3$  dB frequency response of each mixer. The absolute power level accuracy of each waveguide mixer/spectrum analyzer system can be calibrated to within 3 dB by adjusting the front panel AMPLCAL so the display amplitude of a known level external input signal to the mixer is correct.

### OPERATIONAL PRECAUTIONS

1. **Measurements Outside the Specified Frequency and Tuning Range versus Span of the Display.** Signal level or frequency measurements of signals outside the specified frequency range of the band are not reliable.

The total span of the display for some bands exceeds the frequency range; for example, the display extends below the 1.7 GHz lower limit of Band 2 and below the 15 GHz limit of Band 5 (15–21 GHz). The center frequency tuning range and the frequency indicating dot correspond to the specified frequency range of the band. Because of this difference, it can be confusing when the frequency dot fails to tune across a full MAX span display, or a displayed signal outside the frequency range of the band will not tune to center screen. This occurs in Bands 1 and 2 of Option 1 instruments and in Band 5 (15–21 GHz) of all instruments.

The frequency range for external mixer bands is commensurate with the frequency range of the Waveguide Mixers. The span of the displays, however, for these external mixer bands, is much wider. Band 6 (18–26.5 GHz) for example, displays at least a 15–39 GHz span.

2. **Signal FM.** Check to see if the PHASE LOCK switch is activated before deciding that the displayed signal is FMing or the spectrum analyzer is malfunctioning.

3. **Correct Trigger Mode.** The triggering mode is usually in FREE RUN. In pulsed RF applications, a triggered display is required to measure between pulse repetition lines for determining the pulse repetition rate.

Since INTERNAL triggering requires one or more division of signal amplitude, tune the center frequency so a reasonable sized signal is located at the sweep start before changing the trigger source from FREE RUN to INT.

4. **Level of Pulsed Signals.** The spectrum for a pulsed signal is spread out. Consequently, the height of the crt response is less for a pulsed signal than for a cw signal of

the same peak amplitude. This loss in display height means in effect a loss in sensitivity. The amount of loss can be computed from

$$\text{voltage loss} = (t_0 B)^{-1} \text{ where}$$

$$t_0 = \text{pulse}$$

$$B = \text{resolution bandwidth}$$

The spectrum analyzer self-generated noise power increase is proportional to bandwidth. Pulsed RF voltage level is also proportional. Since power is proportional to voltage squared, a wider bandwidth gives better sensitivity and greater dynamic range for pulsed RF inputs.

When in doubt about signal level overdrive problems, reduce the signal level by inserting RF attenuation, then repeat the measurement. If the two agree, the measurement is correct; if not, the input mixer stage is probably overdriven.

An important consideration for pulsed RF measurements is the peak signal level at the mixer. It is greater by  $(t_0 B)^{-1}$  than the peak level displayed on the crt. Taking the sensitivity loss into account is the only sure way of ascertaining that the mixer peak power input for linear operation is not exceeded.

**5. Level of Continuous Wave Signals.** Similar problems can occur when analyzing cw signals at relatively narrow span widths. The large cw signal may not appear on screen because its frequency is outside the set span width. The mixer nevertheless is saturated and will compress signals.

**6. Excessive Input Signal Level.** Too much input power will destroy the front end mixer or attenuator. Replacement mixers and attenuators are costly. When working with high power signals, use couplers or other devices to reduce the signal down to acceptable levels. Once the signal is down below the rating of the RF attenuator, prevent possible mixer damage by starting with the MIN RF ATTEN fully in, then reduce attenuation if needed.

**7. No Crt Trace.** The BASELINE CLIP is used to reduce the intensity of the baseline. If Triggering, Intensity, Vertical Position, etc., all seem to be in order and there is no crt trace, check the BASELINE CLIP state.

**8. Digital Storage Effects on Signal Analyses.** When operating with digital storage, the frequency base is divided into storage slots. For peak displays (above the PEAK/AVERAGE cursor) the display point in each slot corresponds to the maximum sampled value of the signal. Samples are taken at about  $9 \mu\text{s}$  intervals. When sweeping

at one second per division this is about 1000 samples per slot. For average displays (below the cursor) the values of all samples per slot are summed and divided by the number of samples to compute the display point for each slot. Each display point is interconnected to create a smooth display. When A or B are displayed independently, only half of the slots are interconnected. The following are a few pitfalls that can occur.

For wide spans and relatively narrow resolution bandwidth (50 or more resolutions per division), the resolution bandwidth equals a digital storage slot. If that slot is in A memory and only B memory is displayed, that point of the signal will not be displayed and an erroneous level would result. SAVE A will display the correct value because an algorithm chooses the larger of adjacent display points to store in A memory.

If the PEAK/AVERAGE cursor is set above the signal level, the average value for each digital slot will be displayed. With narrow resolution bandwidths compared to the slot width, the average value of the resolution response shape will be displayed, which has nothing to do with signal amplitude.

To avoid the above pitfalls, it is best to run digital storage with A and B interlaced. Do not set the PEAK/AVERAGE cursor to average a cw signal. It is best to set the cursor about 1/4 division above the signal to be averaged and about 1/2 division below the signal to be analyzed.

None of these restrictions apply when the resolution bandwidth is wide compared to a digital storage slot (e.g., 50 MHz/div with 1 MHz resolution).

**9. Stored Display Averaged in Wide Spans.** When operating in wide spans, with digital storage, low level signals will be averaged with the noise and lost if the PEAK/AVERAGE cursor is above the display. Turn the control fully counterclockwise for peak detection when operating with wide spans.

**10. Cold Storage or Power-Interrupt Initialization.** After storage below the operating temperature range (see Environmental Characteristics under Specification), the microcomputer may not power up correctly. If so, allow the instrument to warm up for at least 15 minutes and reinitialize the microcomputer; turn power off for five seconds, then turn it back on. Repeat, if necessary. It may also be necessary to reinitialize the microcomputer after a power interruption.



### Service Manual

The 492 Service Instruction manuals are separate publications that include circuit description, troubleshooting information, calibration procedures, schematic diagrams, and other maintenance information. Service manuals are intended for use by QUALIFIED SERVICE PERSONNEL ONLY. To avoid electrical shock, DO NOT perform any servicing unless qualified to do so.

### Product Service

To assure adequate product service and maintenance for our instruments, Tektronix, Inc. has established Field Offices and Service Centers at strategic points throughout the United States and outside the United States in all countries where our products are sold. Contact your local Service Center, representative or sales engineer for details regarding: Warranty, Calibration, Emergency Repair, Repair Parts, Scheduled Maintenance, Maintenance Agreements, Pickup and Delivery, On-Site Service for fixed installations, and other services available through these centers.

**Emergency Repair:** This service provides immediate attention to instrument malfunction if you are in an emergency situation such as a field trip. Again, contact any Tektronix Service Center for assistance to get you on your way within a minimum of time.

**Maintenance Agreements:** Your instrument is initially covered by warranty (see Warranty statement on the inside of the title page). After the warranty period, several types of maintenance or repair agreements are available. For example: for a fixed fee, a maintenance agreement program provides maintenance and recalibration on a regular basis. Tektronix, Inc. will remind you when a product is due for recalibration and perform the service within a specified time-frame. Any Service Center can furnish complete information on costs and types of maintenance programs.

# OPTION INFORMATION

Information on or about options for the 492 is integrated into the appropriate sections of the manual and clearly identified. At this bind, the following options are available and information is provided in the text:

Option 1: Adds preselector ahead of the 1st mixer, see Section 1.

Option 2: Adds digital storage, see Section 1.

Option 3: Adds 1st LO phase lock, see Section 1.

Option 8: Removes External Mixer capability, see Section 1.

Option 20: Includes General Purpose Waveguide Mixers, 12.4—40.0 GHz, see Section 1.

Option 21: Includes High Performance Waveguide Mixers, 18—40 GHz, see Section 1.

Option 22: Includes High Performance Waveguide Mixers, 18—60 GHz, see Section 1.

Options for power cord configuration, see Section 1.

# GLOSSARY

The following glossary is presented as an aid to better understand the terms as they are used in this document.

## General Terms

**Baseline Clipper (Intensifier).** Increasing the brightness of the signal relative to the baseline portion of the display.

**Center Frequency.** That frequency which corresponds to the center of a frequency span, expressed in hertz.

**Effective Frequency Range.** That range of frequency over which the instrument performance is specified. The lower and upper limits are expressed in hertz.

**Envelope Display.** The display produced on a spectrum analyzer when the resolution bandwidth is greater than the spacing of the individual frequency components.

**Frequency Band.** A part of effective frequency range over which the frequency can be adjusted, expressed in hertz.

**Full Span (Maximum Span).** A mode of operation in which the spectrum analyzer scans an entire frequency band.

**Intermodulation Spurious Response (Intermodulation Distortion).** An unwanted spectrum analyzer response resulting from the mixing of the  $n$ th order frequencies, due to non-linear elements of the spectrum analyzer, the resultant unwanted response being displayed.

**Line Display.** The display produced on a spectrum analyzer when the resolution bandwidth is less than the spacing of the signal amplitudes of the individual frequency components.

**Line Spectrum.** A spectrum composed of signal amplitudes of the discrete frequency components.

## Maximum Safe Input Power

**WITHOUT DAMAGE.** The maximum power applied at the input which will not cause degradation of the instrument characteristics.

**WITH DAMAGE.** The minimum power applied at the input which will damage the instrument.

**Pulse Stretcher.** A pulse shaper that produces an output pulse, whose duration is greater than that of the input pulse, and whose amplitude is proportional to that of the peak amplitude of the input pulse.

**Scanning Velocity.** Frequency span divided by sweep time and expressed in hertz per second.

**Signal Identifier.** A means to identify the spectrum of the input signal when spurious responses are possible.

**Spectrum Analyzer.** An apparatus which is generally used to display the power distribution of an incoming signal as a function of frequency.

## NOTE

*It is useful in analyzing the characteristics of repetitive electrical waveforms in general, since repetitively sweeping through the frequency range of interest will display all components of the signal.*

**Video Filter.** A post detection low-pass filter.

**Zero Span.** A mode of operation in which the frequency span is reduced to zero.

## Terms Related to Frequency

**Display Frequency.** The input frequency as indicated by the spectrum analyzer and expressed in hertz.

**Frequency Drift.** Gradual shift or change in displayed frequency over the specified time due to internal changes in the spectrum analyzer, and expressed in hertz per second, where other conditions remain constant.

**Frequency Linearity Error.** The error of the relationship between the frequency of the input signal and the frequency displayed (expressed as a ratio).

**Frequency Span (Dispersion).** The magnitude of the frequency band displayed, expressed in hertz or hertz per division.

**Impulse Bandwidth.** The displayed spectral level of an applied pulse divided by its spectral voltage density level assumed to be flat within the pass-band.

**Residual FM (Incidental FM).** Short term displayed frequency instability or jitter due to instability in the spectrum analyzer local oscillators, given in terms of peak-to-peak frequency deviation and expressed in hertz or percent of the displayed frequency.

**Shape Factor (Skirt Selectivity).** The ratio of the frequency separation of the two (60 dB/6 dB) down points on the response curve to the static resolution bandwidth.

**Static (Amplifier) Resolution Bandwidth.** The specified bandwidth of the spectrum analyzer's response to a cw signal, if sweep time is kept substantially long.

**NOTE**

*This bandwidth is the frequency separation of two down points, usually 6 dB, on the response curve, if it is measured either by manual scan (true static method) or by using a very low speed sweep (quasi-static method).*

**Zero Pip (Response).** An output indication which corresponds to zero input frequency.

**Terms Related to Amplitude**

**Deflection Coefficient.** The ratio of the input signal magnitude to the resultant output indication.

**NOTE**

*The ratio may be expressed in terms of volts (rms) per division, decibels per division, watts per division, or any other specified factor.*

**Display Dynamic Range.** The maximum ratio of the levels of two non-harmonically related sinusoidal signals each of which can be simultaneously measured on the screen to a specified accuracy.

**Display Flatness.** The unwanted variation of the displayed amplitude over a specified frequency span, expressed in decibels.

**Display Law.** The mathematical law that defines the input-output function of the instrument.

**NOTE**

*The following cases apply:*

1. *Linear—A display in which the scale divisions are a linear function of the input signal voltage.*
2. *Square law (power)—A display in which the scale divisions are a linear function of the input signal power.*
3. *Logarithmic—A display in which the scale divisions are a logarithmic function of the input signal voltage.*

**Display Reference Level.** A designated vertical position representing a specified input level.

**NOTE**

*The level may be expressed in decibels (e.g., 1 mW), volts, or any other units.*

**Dynamic Range.** The maximum ratio of the levels of two signals simultaneously present at the input which can be measured to a specified accuracy.

**Equivalent Input Noise Sensitivity.** The average level of a spectrum analyzer's internally generated noise referenced to the input.

**Frequency Response.** The unwanted variation of the displayed amplitude over a specified center frequency range, measured at the center frequency, expressed in decibels.

**Gain Compression.** Maximum input level where the scale linearity error is below that specified.

**Hum Sidebands.** Undesired responses created within the spectrum analyzer, appearing on the display that are separated from the desired response by the fundamental or harmonic of the power-line frequency.

**Input Impedance.** The impedance at the desired input terminal.

**NOTE**

*Usually expressed in terms of VSWR, return loss, or other related terms for low impedance devices and resistance-capacitance parameters for high impedance devices.*

**Noise Sidebands.** Undesired response caused by noise internal to the spectrum analyzer appearing on the display around a desired response.

**Relative Display Flatness.** The display flatness measured relative to the display amplitude at a fixed frequency within the frequency span, expressed in decibels.

**NOTE**

*Display flatness is closely related to frequency response. The main difference is that the spectrum display is not recentered.*

**Residual Response.** A spurious response in the absence of an input signal. (Noise and zero pip are excluded.)

**Sensitivity.** Measure of a spectrum analyzer's ability to display minimum level signals, at a given IF bandwidth, display mode, and any other influencing factors, and expressed in decibels (e.g., 1 mW).

**Spurious Response.** A response of a spectrum analyzer wherein the displayed frequency does not conform to the input frequency.

## Terms Related to Digital Storage for Spectrum Analyzers

**Clear (Erase).** Presets memory to a prescribed state, usually that denoting zero.

**Digitally Averaged Display.** A display of the average value of digitized data computed by combining serial samples in a defined manner.

**Digitally Stored Display.** A display method whereby the displayed function is held in a digital memory. The display is generated by reading the data out of memory.

**Max Hold (Peak Mode).** Digitally stored display mode which, at each frequency address, compares the incoming signal level to the stored level and retains the greater. In this mode, the display indicates the peak level at each frequency after several successive sweeps.

**Multiple Display Memory.** A digitally stored display having multiple memory sections which can be displayed separately or simultaneously.

**Save.** A function which inhibits storage update, saving existing data in a section of a multiple memory (e.g., Save A).

**Scan Address.** A number representing each horizontal data position increment on a directed beam type display. An address in a memory is associated with each scan address.

**View (Display).** Enables viewing of contents of the chosen memory section (e.g., "View A" displays contents of memory A; "View B" displays the contents of memory B).

## MANUAL CHANGE INFORMATION

At Tektronix, we continually strive to keep up with latest electronic developments by adding circuit and component improvements to our instruments as soon as they are developed and tested.

Sometimes, due to printing and shipping requirements, we can't get these changes immediately into printed manuals. Hence, your manual may contain new change information on following pages.

A single change may affect several sections. Since the change information sheets are carried in the manual until all changes are permanently entered, some duplication may occur. If no such change pages appear following this page, your manual is correct as printed.

## SERVICE NOTE

Because of the universal parts procurement problem, some electrical parts in your instrument may be different from those described in the Replaceable Electrical Parts List. The parts used will in no way alter or compromise the performance or reliability of this instrument. They are installed when necessary to ensure prompt delivery to the customer. Order replacement parts from the Replaceable Electrical Parts List.

# CALIBRATION TEST EQUIPMENT REPLACEMENT

## Calibration Test Equipment Chart

This chart compares TM 500 product performance to that of older Tektronix equipment. Only those characteristics where significant specification differences occur, are listed. In some cases the new instrument may not be a total functional replacement. Additional support instrumentation may be needed or a change in calibration procedure may be necessary.

### Comparison of Main Characteristics

DM 501 replaces 7D13		
PG 501 replaces 107 108	PG 501 - Risetime less than 3.5 ns into 50 $\Omega$ . PG 501 - 5 V output pulse; 3.5 ns Risetime	107 - Risetime less than 3.0 ns into 50 $\Omega$ . 108 - 10 V output pulse 1 ns Risetime
PG 502 replaces 107 108 111	PG 502 - 5 V output PG 502 - Risetime less than 1 ns; 10 ns Pretrigger pulse delay	108 - 10 V output 111 - Risetime 0.5 ns; 30 to 250 ns Pretrigger pulse delay
PG 508 replaces 114 115 2101	Performance of replacement equipment is the same or better than equipment being replaced.	
PG 506 replaces 106 067-0502-01	PG 506 - Positive-going trigger output signal at least 1 V; High Amplitude output, 60 V. PG 506 - Does not have chopped feature.	106 - Positive and Negative-going trigger output signal, 50 ns and 1 V; High Amplitude output, 100 V. 0502-01 - Comparator output can be alternately chopped to a reference voltage.
SG 503 replaces 190, 190A, 190B 191 067-0532-01	SG 503 - Amplitude range 5 mV to 5.5 V p-p. SG 503 - Frequency range 250 kHz to 250 MHz.	190B - Amplitude range 40 mV to 10 V p-p. 0532-01 - Frequency range 65 MHz to 500 MHz.
SG 504 replaces 067-0532-01 067-0650-00	SG 504 - Frequency range 245 MHz to 1050 MHz.	0532-01 - Frequency range 65 MHz to 500 MHz.
TG 501 replaces 180, 180A 181 184 2901	TG 501 - Trigger output-slaved to marker output from 5 sec through 100 ns. One time-mark can be generated at a time. TG 501 - Trigger output-slaved to market output from 5 sec through 100 ns. One time-mark can be generated at a time. TG 501 - Trigger output-slaved to marker output from 5 sec through 100 ns. One time-mark can be generated at a time.	180A - Trigger pulses 1, 10, 100 Hz; 1, 10, and 100 kHz. Multiple time-marks can be generated simultaneously. 181 - Multiple time-marks 184 - Separate trigger pulses of 1 and 0.1 sec; 10, 1, and 0.1 ms; 10 and 1 $\mu$ s. 2901 - Separate trigger pulses, from 5 sec to 0.1 $\mu$ s. Multiple time-marks can be generated simultaneously.

NOTE: All TM 500 generator outputs are short-proof. All TM 500 plug-in instruments require TM 500-Series Power Module.

## DESCRIPTION

## TEXT CORRECTIONS

## SECTION 1

Specification - Electrical Characterisitics

Page 1-5

Amplitude Related - IF Gain Accuracy:

## CHANGE TO READ:

Within 0.2 dB/dB except at the 10 dB transition from -29 to -30 dBm, -39 to -40 dBm, -49 to -50 dBm, -59 to -60 dBm, and -60 to -70 dBm. where the accuracy is 0.5 dB; and within 0.5 dB/10 dB to a maxium of  $\pm 2$  dB over the 90 dB range.



### DESCRIPTION

#### TEXT CHANGES

#### SECTION 1 GENERAL INFORMATION AND SPECIFICATIONS

Page 1-4 ELECTRICAL CHARACTERISTICS(cont), Table 1-1

CHANGE the LO Frequency & Harmonic (n) for the following Frequency Bands:

Frequency Band	LO Frequency & Harmonic (n)
4	(2076-6276) (3-)
6	(2655-3988) (6+)
10	(3823-5999) (23+)

Page 1-7 ELECTRICAL CHARACTERISTICS

CHANGE the following Supplemental Information:

Characteristics	Supplemental Information
External Mixer	Input for IF signal and the source of negative-going bias for external waveguide mixers. Bias range +1.0 to -2.0 V, source impedance 70 ohms.

Page 1-12 Table 1-4, OPTION 1 FREQUENCY CHARACTERISTICS

CHANGE the LO Frequency & Harmonic (n) for the following Frequency Bands:

Frequency Band	LO Frequency & Harmonic (n)
4	(2076-6276) (3-)
6	(2655-3988) (6+)
10	(3823-5999) (23+)

Page 1-17 Fig. 1-2

CHANGE A2 current rating to 13A.

#### SECTION 3 OPERATION

Page 3-2 Item 10

ADD after last sentence:

Range of tuning is limited to 4.5 MHz in the narrow spans because a second oscillator is switched to improve stability.

## DESCRIPTION

Page 3-7 Figures 3-3 and 3-4

REPLACE with the following:




Fig. 3-3. CRT readout for power-up state (Option 1).




Fig. 3-4. Typical display of calibrator markers in MAX SPAN position.

## DESCRIPTION

Page 3-17 Step 8, part a.

CHANGE first sentence to read:

a. Set the FREQUENCY RANGE to Band 1 (0-4.1 GHz or 0-1.8 GHz Option 1).

Page 3-19

GENERAL OPERATING INFORMATION, RF INPUT Connector, 2nd paragraph

ADD:

Line stabilizing networks used for conducting EMI/RF measurements, will often have several volts of 60 Hz signal at the output. A dc block, such as Tektronix part 015-0221-00, will protect the input mixer and prevent destruction.

Page 3-25

Second paragraph

CHANGE to read:

.....of LO signal with a variable bias from -2.0 V to +1.0 V through a current limiting resistor.....

### DESCRIPTION

#### TEXT CORRECTIONS

#### SECTION 1 SPECIFICATION

Page 1-10 PHYSICAL CHARACTERISTICS, Description

#### CHANGE TO READ:

With front cover, feet and handle 9.15 X 15.03 X 23.1 inches (handle folded back over instrument), 25.85 inches (handle fully extended).

Page 1-11 Table 1-3

ADD to Diplexer Assembly

Adapter; TNC-to-SMA	015-0388-00	1	Part of Diplexer Assembly
---------------------	-------------	---	---------------------------

Cable; semi-rigid, 50 ohm	015-1055-00	1	Part of Diplexer Assembly
---------------------------	-------------	---	---------------------------

Page 1-12 OPTIONAL ACCESSORIES

DELETE: Circuit Board Extender Assembly	672-0865-00	1
---	-------------	---

Frame (extrusion) for Circuit Board Extender	426-1527-00	1
--	-------------	---

Left Extender Board	670-5562-00	1
---------------------	-------------	---

Right Extender Board	670-5563-00	2
----------------------	-------------	---

Screws, panhead with flat and lockwashers	211-0116-00	6
---	-------------	---

Page 3-3 Table 3-1, SPAN/DIV RANGES VERSUS BAND AND OPTION

CHANGE Standard column to read:

10 kHz

50 kHz

50 kHz

100 kHz

200 kHz

500 kHz

500 kHz

Page 3-6

④ VERT

ADD:

If signal is used to drive a chart recorder, digital storage should be off. If signal drives a slave monitor and PEN LIFT is used for blanking, the screen will blank during retrace time.

## DESCRIPTION

Page 3-19 GENERAL OPERATING INFORMATION, RF INPUT Connector, 2nd paragraph  
ADD: Line stabilizing networks used for conducting EMI/RF measurements, will often have several volts of 60 Hz signal at the output. A dc block will protect the input mixer and prevent destruction.

Page 3-24 WAVEGUIDE MIXERS AND EXTERNAL DIPLEXER, Installation, paragraph one  
ADD: The Diplexer assembly includes a SMA-to-TNC adapter and a shaped semi-rigid coaxial cable. Cable length is not critical, therefore the semi-rigid cable can be replaced with different length cable.

Date: 3-4-80 Change Reference: C3/380  
Product: 492 SPECTRUM ANALYZER OPERATORS Manual Part No.: 070-2726-01

DESCRIPTION

TEXT CORRECTIONS

Page 1-11 TABLE 1-3

ACCESSORIES

DELETE:

Crt Light Filter, Blue	378-0115-03	1	Front Cover
------------------------	-------------	---	-------------

CHANGE:

From:  
Crt Mesh Filter 378-0726-01

To:

Crt Mesh Filter 378-0887-00

### DESCRIPTION

#### TEXT CORRECTION

Section 1, Option 1, Specifications, page 1-14

CHANGE the Performance Requirement to read:

Characteristic	Performance Requirement
LO emission, referenced to input mixer and with zero RF attenuation	$\leq -70$ dBm to 18 GHz

## DESCRIPTION

## 492P Programmable Spectrum Analyzer

This manual also applies to local operation of the TEKTRONIX 492P Programmable Spectrum Analyzer. For information on controls available on the 492P, but not the 492, see Section 1 of the 492P Programmer's Manual, which covers topics such as the 492P RESET TO LOCAL button, GPIB ADDRESS switches, and talk/listen-only modes.

Page 1-2

Insert in the table of Electrical Characteristics between the entries for Accuracy and Readout Resolution:

TUNE command accuracy (after a 2 hour warmup--applies only to 492P under remote control) (+/-7% or +/-150 kHz) \* n, whichever is greater; see Table 1-1 for the value of n in a given band.

Repeatability of frequency setting (ambient temperature change must be 10 degrees C--applies only to 492P under remote control)

+/- (2 MHz +10% of span/div) or +/- (0.1% of frequency + 10% of span/div) of previous settings to the same frequency, whichever is greater.



## DESCRIPTION

## TEXT CORRECTIONS

SECTION 3 OPERATION

Page 3-7 Insert the following prior to TURN ON PROCEDURE:

## Firmware Version

During initial turn-on or power-up cycle, the firmware version in the instrument will flash on screen for approximately two seconds. The Replacement Parts List in the Service manual lists the ROM's used for each version.

## Error Message Readout

The 492/492P features error message readout. These messages (numbers) flash on screen when the address and data from the microcomputer fails to complete an operational routine. These error numbers and their meaning are as follows:

<u>Error #</u>	<u>Meaning</u>
57	Tune routine failed
58	Failed to phase lock
59	Lost phase lock
60	Failed to recenter frequency when phase lock cancelled or when switching to an unlocked span/div setting. (Phase lock occurs for 50 kHz or less in bands 1 through 3, 100 kHz or less for band 4, and 200 kHz or less for bands 5 and above.)

DESCRIPTION

Page 3-15 Table 3-3

Change Option 3, 100 Hz sensitivities to read, from top to bottom:

- 123
- 108
- 103
- 108
- 103
- 103

Page 3-16 Table 3-4

Change Option 3, 100 Hz sensitivities to read, from top to bottom:

- 118
- 103
- 93
- 108
- 103
- 103

Page 3-16

Step 7 Frequency Drift or Stability, part c.

Add:

Allow a period of at least 30 minutes at a fixed frequency before performing the drift check.

### DESCRIPTION

#### TEXT CORRECTION

SECTION 1      Specification

Page 1-7      Input Level Characteristic, 1 dB Compression point

CHANGE TO READ AS FOLLOWS AND MOVE TO SUPPLEMENTAL COLUMN

<u>Characteristics</u>	<u>Performance Requirements</u>	<u>Supplemental Information</u>
1 dB compression point		-18 dBm, no RF attenuation.



050-1569-01

M39979, M60460

## HANDLE REPLACEMENT

For the following TEKTRONIX<sup>®</sup> Instruments:

492/P      Serial Numbers B010100 - B041979  
496/P      Serial Numbers B010100 - B010279

This Parts Replacement kit contains parts and instructions to replace the carrying handle with an improved carrying handle that is better able to support the weight of the instrument.

### NOTE

If the serial number of your instrument is greater than those listed above, or this kit has been installed, disregard the instructions and use any of the parts, provided in the kit, as a direct replacements.

KIT PARTS LIST:

Quantity	Part Number	Description
1 ea	-----	HANDLE ASSEMBLY CONSISTING OF:
2 ea	105-0893-00	INDEX, RELEASE, HANDLE
2 ea	105-0894-00	INDEX STOP, HANDLE
2 ea	200-2622-00	CAP, HANDLE
2 ea	200-2694-00	COVER, INDEX
6 ea	212-0560-00	SCREW, MACHINE, 10-32 x 0.312 FLH
6 ea	213-0918-00	SCREW, 8-32 x 0.625 FLH
2 ea	214-3198-00	INDEX, HANDLE
2 ea	214-3199-01	BUTTON, RELEASE
2 ea	214-3200-00	SPRING, HLCPS
2 ea	334-4661-00	MARKER, IDENTIFICATION, "PUSH"
1 ea	367-0306-01	HANDLE, CARRYING
2 ea	380-0668-00	HOUSING, INDEX

INSTRUCTIONS:

WARNING

Before proceeding, ensure the mainframe power switch is in the off position, then disconnect the instrument from the power source.

- ( ) 1. Place the instrument on its face or front panel, loosen the four cabinet mounting screws and remove the cabinet.
- ( ) 2. Remove the eight (four each side) screws that fasten the carrying handle to the cabinet and remove the carrying handle.
- ( ) 3. Place the handle assembly (provided in the kit) around the open (front) end of the cabinet, with the vented side of the cabinet down. The word "TEKTRONIX" on the handle (facing the front of the cabinet) should be right side up.
- ( ) 4. Install and tighten the eight screws removed in step 2.
- ( ) 5. Reinstall the cabinet and tighten the cabinet mounting screws.
- ( ) 6. For future reference, fasten the Manual Modification Insert in your Service Manual, Volume 2.

PRODUCT MODIFICATION KIT  
SUGGESTION/CORRECTION FORM

DATE \_\_\_\_\_

KIT NUMBER \_\_\_\_\_ STEP/PAGE \_\_\_\_\_

FIGURE NUMBER \_\_\_\_\_ PUBLICATION DATE \_\_\_\_\_

DISCREPANCY \_\_\_\_\_

SUGGESTED CORRECTION/COMMENTS \_\_\_\_\_

SUGGESTED BY: NAME/ORGANIZATION \_\_\_\_\_

\_\_\_\_ REPLY REQUESTED

(PLEASE TYPE OR PRINT LEGIBLY)

RETURN TO LOCAL FIELD OFFICE/SERVICE CENTER

FIELD OFFICE/SERVICE CENTER/DEL. STA. \_\_\_\_\_

SERVICE CENTER: RETURN TO FIELD MOOS 78-587

---

REPLY

\_\_\_\_ WILL MAKE CHANGE IMMEDIATELY

\_\_\_\_ WILL MAKE CHANGE AT NEXT PRINTING

\_\_\_\_ OTHER \_\_\_\_\_

SIGNED \_\_\_\_\_ DATE \_\_\_\_\_

---

# TEKTRONIX

## MANUAL MODIFICATION INSERT

---

### HANDLE REPLACEMENT

for

492/P            Serial Number B010100 - B041979  
496/P            Serial Number B010100 - B010279

Installed in SN \_\_\_\_\_ Date \_\_\_\_\_

This modification insert is provided to supplement the manual for the above listed product(s). The information given in this insert supersedes that given in the manual.

Copyright © 1986 by Tektronix, Inc., Beaverton, Oregon. Printed in the United States of America. All rights reserved. Contents of this insert may not be reproduced in any form without the permission of the copyright owner.

### GENERAL INFORMATION

This Parts Replacement kit contains parts and instructions to replace the carrying handle with an improved carrying handle that is better able to support the weight of the instrument.

## REPLACEABLE MECHANICAL PARTS

Index No.	Part Number	Quantity	Description
1	213-0918-00	6 ea	SCREW, MACH. 8-32 x 0.625 FLH.
2	200-2694-00	1 ea	COVER, INDEX
3	214-3200-00	1 ea	SPRING, HLCPS
4	105-0893-00	1 ea	INDEX RELEASE, HANDLE
5	105-0894-00	1 ea	INDEX STOP, HANDLE
6	214-3198-00	1 ea	INDEX, HANDLE
7	380-0668-00	1 ea	HOUSING, INDEX
8	213-0858-00	8 ea	SCREW, MACH. 6-32 x 0.500 FLH.
9	367-0306-00 <sup>01</sup>	1 ea	HANDLE, CARRYING
10	214-3199-01	2 ea	BUTTON, RELEASE
11	334-4661-00	2 ea	CAP, HANDLE
12	212-0560-00	6 ea	SCREW, MACH: 10-32 X 0.312 FLH
13	200-2622-00	2 ea	CAP, HANDLE






Figure 1. Partial - cabinet